

LB129170


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

50863



A= 0.0000e+000

B= 1.9911e-004

C= -7.3290e-002

Rho= 0.9998992

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	o/b)
0.0	0.000	0.012	0.012	429	0.000	429	0				10
0.2	0.200	0.180	-0.020	1270	0.0 %	1270					1
2.5	2.500	2.524	0.024	13045	0.0 %	13045					1
5.0	5.000	5.036	0.036	25659	0.0 %	25659					1
7.5	7.500	7.395	-0.105	37506	0.0 %	37506					1
10.0	10.000	10.054	0.054	50863	0.0 %	50863					1

LB129170 INSTRUMENT ID:CV1

Method:7470A

Date Of Analysis : 18 jan 2024

Sample ID	Extended ID	μ Abs.	Conc.	Stnd Conc	Method	Units	Date	Type
	0	429	-	0	7470A	PPB	1/18/2024 11:44	Std
	0.2	1270	-	0.2	7470A	PPB	1/18/2024 11:47	Std
	2.5	13045	-	2.5	7470A	PPB	1/18/2024 11:49	Std
	5	25659	-	5	7470A	PPB	1/18/2024 11:51	Std
	7.5	37506	-	7.5	7470A	PPB	1/18/2024 11:54	Std
	10	50863	-	10	7470A	PPB	1/18/2024 11:59	Std
ICV70	ICV70	19691	3.8474 -		7470A	PPB	1/18/2024 12:02	SMPL
ICB70	ICB70	153	-0.0428 -		7470A	PPB	1/18/2024 12:04	SMPL
CCV85	CCV85	25688	5.0414 -		7470A	PPB	1/18/2024 12:06	SMPL
CCB85	CCB85	119	-0.0496 -		7470A	PPB	1/18/2024 12:09	SMPL
HighStd	HighStd	50820	10.0455 -		7470A	PPB	1/18/2024 12:13	SMPL
ChkStd	ChkStd	36046	7.1038 -		7470A	PPB	1/18/2024 12:15	SMPL
CRA	CRA	1252	0.176 -		7470A	PPB	1/18/2024 12:21	SMPL
PB158487BL	PB158487BL	272	-0.0191 -		7470A	PPB	1/18/2024 12:23	SMPL
PB158487BS	PB158487BS	24280	4.7611 -		7470A	PPB	1/18/2024 12:26	SMPL
P1150-02	MR-306-0-2	392	0.0048 -		7470A	PPB	1/18/2024 12:28	SMPL
P1150-04	MR-306-25-27	632	0.0525 -		7470A	PPB	1/18/2024 12:30	SMPL
P1150-04DUP	MR-306-25-27DUP	400	0.0064 -		7470A	PPB	1/18/2024 12:33	SMPL
P1150-04MS	MR-306-25-27MS	22892	4.4847 -		7470A	PPB	1/18/2024 12:35	SMPL
P1150-04MSD	MR-306-25-27MSD	24229	4.7509 -		7470A	PPB	1/18/2024 12:37	SMPL
CCV86	CCV86	27050	5.3126 -		7470A	PPB	1/18/2024 12:39	SMPL
CCB86	CCB86	87	-0.056 -		7470A	PPB	1/18/2024 12:42	SMPL
PB158446TB	PB158446TB	571	0.0404 -		7470A	PPB	1/18/2024 12:44	SMPL
PB158464TB	PB158464TB	666	0.0593 -		7470A	PPB	1/18/2024 12:46	SMPL
P1150-04LX5	MR-306-25-27L	140	-0.0454 -		7470A	PPB	1/18/2024 12:48	SMPL
P1150-04A	MR-306-25-27A	21913	4.2898 -		7470A	PPB	1/18/2024 12:51	SMPL
CCV87	CCV87	26947	5.2921 -		7470A	PPB	1/18/2024 12:53	SMPL
CCB87	CCB87	94	-0.0546 -		7470A	PPB	1/18/2024 12:55	SMPL