


LB129176

SFAM01.1

INSTRUMENT ID: CV1

Linear



A= 0.0000e+000

B= 1.9750e-004 slope

C= -4.5471e-002 intercept

Rho= 0.9997425

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	O/O
0.00	0.000	-0.033	-0.033	65	0.000	65					=
0.05	0.050					442	0				-10
0.20	0.200	0.180	-0.020	1141	0.0 %	1141					-2
2.50	2.500	2.456	-0.044	12664	0.0 %	12664					3
5.00	5.000	5.150	0.150	26307	0.0 %	26307					-1
7.50	7.500	7.559	0.059	38506	0.0 %	38506					-1
10.0	10.000	9.888	-0.112	50296	0.0 %	50296					

LB129176

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	65	-		0 SFAM01.1	PPB	1/18/2024 15:47	Std
	0.2 S01	1141	-		0.2 SFAM01.1	PPB	1/18/2024 15:49	Std
	2.5 S02	12664	-		2.5 SFAM01.1	PPB	1/18/2024 15:51	Std
	5 S03	26307	-		5 SFAM01.1	PPB	1/18/2024 15:54	Std
	7.5 S04	38506	-		7.5 SFAM01.1	PPB	1/18/2024 15:56	Std
	10 S05	50296	-		10 SFAM01.1	PPB	1/18/2024 15:58	Std
ICV071	ICV071	19490	3.8037 -		SFAM01.1	PPB	1/18/2024 16:01	SMPL
ICB071	ICB071	-106	-0.0664 -		SFAM01.1	PPB	1/18/2024 16:03	SMPL
CCV088	CCV088	24648	4.8224 -		SFAM01.1	PPB	1/18/2024 16:06	SMPL
CCB088	CCB088	-127	-0.0706 -		SFAM01.1	PPB	1/18/2024 16:08	SMPL
PB158512BL	PBW512	167	-0.0125 -		SFAM01.1	PPB	1/18/2024 16:10	SMPL
PB158466TB	LEB466	215	-0.003 -		SFAM01.1	PPB	1/18/2024 16:13	SMPL
P1116-01	EOX68	231	0.0002 -		SFAM01.1	PPB	1/18/2024 16:15	SMPL
P1116-02	EOX68D	212	-0.0036 -		SFAM01.1	PPB	1/18/2024 16:17	SMPL
P1116-03	EOX68S	5865	1.1128 -		SFAM01.1	PPB	1/18/2024 16:19	SMPL
CCV089	CCV089	24311	4.7559 -		SFAM01.1	PPB	1/18/2024 16:22	SMPL
CCB089	CCB089	-168	-0.0787 -		SFAM01.1	PPB	1/18/2024 16:24	SMPL