


7471B

LB129185
INSTRUMENT 412: CV1

Linear

 μ Abs.:

57928



A= 0.0000e+000

B= 1.7318e-004

C= 2.0000e-002

Rho= 0.9999308

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	o/o)
0.0	0.000	0.052	0.052	184	0.000	184					
0.2	0.200	0.211	0.011	1105	0.0 %	1105					
2.5	2.500	2.439	-0.061	13969	0.0 %	13969					
5.0	5.000	4.990	-0.010	28695	0.0 %	28695					
7.5	7.500	7.456	-0.044	42936	0.0 %	42936					
10.0	10.000	10.052	0.052	57928	0.0 %	57928					

o/o)

1

1

0

2

5

1

LB129185 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 19 Jan 2024 13:20:18

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	184	-	0.0000	7471B	PPB	19 Jan 2024 13:24:28	S	Std
0.2 - 1	52	1105	-	0.2000	7471B	PPB	19 Jan 2024 13:26:45	S	Std
2.5 - 1	52.5	13969	-	2.5000	7471B	PPB	19 Jan 2024 13:29:02	S	Std
5.0 - 1	55	28695	-	5.0000	7471B	PPB	19 Jan 2024 13:31:19	S	Std
7.5 - 1	57.5	42936	-	7.5000	7471B	PPB	19 Jan 2024 13:33:38	S	Std
10.0 - 1	58	57928	-	10.0000	7471B	PPB	19 Jan 2024 13:36:03	S	Std
ICV73 - 1	ICV73	23515	4.0924	-	7471B	PPB	19 Jan 2024 13:39:19	U	SMPL
ICB73 - 1	ICB73	-216	-0.0174	-	7471B	PPB	19 Jan 2024 13:41:35	U	SMPL
CCV93 - 1	CCV93	28883	5.0221	-	7471B	PPB	19 Jan 2024 13:43:57	U	SMPL
CCB93 - 1	CCB93	-210	-0.0164	-	7471B	PPB	19 Jan 2024 13:46:12	U	SMPL
CRA - 1	CRA	1059	0.2034	-	7471B	PPB	19 Jan 2024 13:48:35	U	SMPL
HighStd - 1	HighStd	57529	9.9831	-	7471B	PPB	19 Jan 2024 13:50:50	U	SMPL
ChkStd - 1	ChkStd	38665	6.7162	-	7471B	PPB	19 Jan 2024 13:53:07	U	SMPL
PB158513BL - 1	PB158513BL	-315	-0.0346	-	7471B	PPB	19 Jan 2024 13:55:38	U	SMPL
PB158513BS - 1	PB158513BS	25646	4.4615	-	7471B	PPB	19 Jan 2024 13:58:06	U	SMPL
P1150-01 - 1	MR-306-0-2	7705	1.3544	-	7471B	PPB	19 Jan 2024 14:00:23	U	SMPL
P1150-03 - 1	MR-306-25-27	4789	0.8494	-	7471B	PPB	19 Jan 2024 14:02:51	U	SMPL
P1152-03 - 1	275	5919	1.0451	-	7471B	PPB	19 Jan 2024 14:05:10	U	SMPL
P1153-01 - 1	PAD-A	767	0.1528	-	7471B	PPB	19 Jan 2024 14:07:30	U	SMPL
P1153-02 - 1	PAD-B	1301	0.2453	-	7471B	PPB	19 Jan 2024 14:09:49	U	SMPL
CCV94 - 1	CCV94	28520	4.9592	-	7471B	PPB	19 Jan 2024 14:12:07	U	SMPL
CCB94 - 1	CCB94	-58	0.0100	-	7471B	PPB	19 Jan 2024 14:14:22	U	SMPL
P1155-01 - 1	BB-TESTPITS	2422	0.4395	-	7471B	PPB	19 Jan 2024 14:16:45	U	SMPL
P1159-01 - 1	TP-011724	1276	0.2410	-	7471B	PPB	19 Jan 2024 14:19:01	U	SMPL
P1161-01 - 1	SU-02-011724	844	0.1662	-	7471B	PPB	19 Jan 2024 14:21:17	U	SMPL
P1169-01 - 1	WC-A-COMP	21715	3.7807	-	7471B	PPB	19 Jan 2024 14:23:33	U	SMPL
P1169-04 - 1	WC-B-COMP	13955	2.4368	-	7471B	PPB	19 Jan 2024 14:25:49	U	SMPL
P1169-07 - 1	WC-C-COMP	2681	0.4843	-	7471B	PPB	19 Jan 2024 14:28:12	U	SMPL
P1169-07MS - 1	WC-C-COMPMS	30155	5.2424	-	7471B	PPB	19 Jan 2024 14:32:50	U	SMPL
P1169-07MSD - 1	WC-C-COMPMSD	28829	5.0127	-	7471B	PPB	19 Jan 2024 14:35:10	U	SMPL
P1181-01 - 1	TR-05-011824	363	0.0829	-	7471B	PPB	19 Jan 2024 14:37:35	U	SMPL
CCV95 - 1	CCV95	29036	5.0486	-	7471B	PPB	19 Jan 2024 14:40:00	U	SMPL
CCB95 - 1	CCB95	-102	0.0023	-	7471B	PPB	19 Jan 2024 14:42:15	U	SMPL
P1169-07LX5 - 1		346	0.0799	-	7471B	PPB	19 Jan 2024 14:44:39	U	SMPL
P1169-07A - 1		28130	4.8917	-	7471B	PPB	19 Jan 2024 14:46:54	U	SMPL
P11169-07DUP - 1	WC-CCOMPDUP	3201	0.5744	-	7471B	PPB	19 Jan 2024 14:51:51	U	SMPL
CCV96 - 1	CCV96	29434	5.1175	-	7471B	PPB	19 Jan 2024 14:54:07	U	SMPL
CCB96 - 1	CCB96	-49	0.0115	-	7471B	PPB	19 Jan 2024 14:56:24	U	SMPL