

LB129191


SFAM01.1

INSTRUMENT ID: CV1

Linear

μ Abs.:

51125



A= 0.0000e+000

B= 1.9509e-004 slope

C= -2.3711e-002 y-intercept

Rho= 0.9998995

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	total
0.00	0.000	0.005	0.005	149	0.000	149					=
0.05	0.050					442	0				12
0.20	0.200	0.223	0.023	1266	0.0 %	1266					-3
2.50	2.500	2.414	-0.086	12493	0.0 %	12493					1
5.00	5.000	5.046	0.046	25985	0.0 %	25985					1
7.50	7.500	7.562	0.062	38882	0.0 %	38882					1
10.0	10.000	9.950	-0.050	51125	0.0 %	51125					-1

LB129191
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	149	-		0 SFAM01.1	PPB	1/19/2024 15:49	Std
	0.2 S01	1266	-		0.2 SFAM01.1	PPB	1/19/2024 15:51	Std
	2.5 S02	12493	-		2.5 SFAM01.1	PPB	1/19/2024 15:54	Std
	5 S03	25985	-		5 SFAM01.1	PPB	1/19/2024 15:56	Std
	7.5 S04	38882	-		7.5 SFAM01.1	PPB	1/19/2024 15:58	Std
	10 S05	51125	-		10 SFAM01.1	PPB	1/19/2024 16:01	Std
ICV074	ICV074	20617	3.9985 -		SFAM01.1	PPB	1/19/2024 16:04	SMPL
ICB074	ICB074	-122	-0.0475 -		SFAM01.1	PPB	1/19/2024 16:06	SMPL
CCV001	CCV001	26339	5.1148 -		SFAM01.1	PPB	1/19/2024 16:08	SMPL
CCB001	CCB001	-161	-0.0551 -		SFAM01.1	PPB	1/19/2024 16:10	SMPL
PB158533BL	PBW533	54	-0.0132 -		SFAM01.1	PPB	1/19/2024 16:13	SMPL
P1060-01	YBRR6	1095	0.1899 -		SFAM01.1	PPB	1/19/2024 16:15	SMPL
P1060-02	YBRR6D	1177	0.2059 -		SFAM01.1	PPB	1/19/2024 16:17	SMPL
P1060-03	YBRR6S	7264	1.3934 -		SFAM01.1	PPB	1/19/2024 16:20	SMPL
P1060-04	YBRR7	641	0.1013 -		SFAM01.1	PPB	1/19/2024 16:22	SMPL
P1060-05	YBRR8	794	0.1312 -		SFAM01.1	PPB	1/19/2024 16:24	SMPL
P1064-01	EOWX5	17	-0.0204 -		SFAM01.1	PPB	1/19/2024 16:26	SMPL
P1064-02	EOWY5	92	-0.0058 -		SFAM01.1	PPB	1/19/2024 16:29	SMPL
P1064-03	EOWZ3	66	-0.0108 -		SFAM01.1	PPB	1/19/2024 16:31	SMPL
P1064-04	E0X01	68	-0.0104 -		SFAM01.1	PPB	1/19/2024 16:33	SMPL
P1064-05	EOWX7	54	-0.0132 -		SFAM01.1	PPB	1/19/2024 16:35	SMPL
P1064-06	E0X00	53	-0.0134 -		SFAM01.1	PPB	1/19/2024 16:38	SMPL
P1064-07	EOWY4	48	-0.0143 -		SFAM01.1	PPB	1/19/2024 16:40	SMPL
P1064-08	EOWZ2	55	-0.013 -		SFAM01.1	PPB	1/19/2024 16:42	SMPL
P1064-09	E0X17	34	-0.0171 -		SFAM01.1	PPB	1/19/2024 16:44	SMPL
P1064-10	E0X37	40	-0.0159 -		SFAM01.1	PPB	1/19/2024 16:47	SMPL
P1064-11	EOWZ8	123	0.0003 -		SFAM01.1	PPB	1/19/2024 16:49	SMPL
P1064-12	EOWZ8D	129	0.0015 -		SFAM01.1	PPB	1/19/2024 16:51	SMPL
P1064-13	EOWZ8S	5162	0.9833 -		SFAM01.1	PPB	1/19/2024 16:54	SMPL
P1064-14	EOWY2	0	-0.0237 -		SFAM01.1	PPB	1/19/2024 16:56	SMPL
CCV002	CCV002	25592	4.969 -		SFAM01.1	PPB	1/19/2024 16:58	SMPL
CCB002	CCB002	-116	-0.0463 -		SFAM01.1	PPB	1/19/2024 17:00	SMPL
P1064-15	E0X18	0	-0.0237 -		SFAM01.1	PPB	1/19/2024 17:03	SMPL
P1064-16	E0X38	15	-0.0208 -		SFAM01.1	PPB	1/19/2024 17:05	SMPL
PB158534BL	PBW534	31	-0.0177 -		SFAM01.1	PPB	1/19/2024 17:07	SMPL
P1063-01	EOWT5	257	0.0264 -		SFAM01.1	PPB	1/19/2024 17:09	SMPL
P1063-02	EOWW3	411	0.0565 -		SFAM01.1	PPB	1/19/2024 17:12	SMPL
P1063-03	EOWX1	70	-0.0101 -		SFAM01.1	PPB	1/19/2024 17:14	SMPL
P1063-04	EOWX4	99	-0.0044 -		SFAM01.1	PPB	1/19/2024 17:16	SMPL
P1063-05	EOWX0	137	0.003 -		SFAM01.1	PPB	1/19/2024 17:19	SMPL
P1063-06	EOWX6	163	0.0081 -		SFAM01.1	PPB	1/19/2024 17:21	SMPL
P1063-07	EOWT4	160	0.0075 -		SFAM01.1	PPB	1/19/2024 17:23	SMPL
P1063-08	EOWW2	154	0.0063 -		SFAM01.1	PPB	1/19/2024 17:25	SMPL
P1063-09	E0X07	109	-0.0024 -		SFAM01.1	PPB	1/19/2024 17:28	SMPL
P1063-10	E0X27	175	0.0104 -		SFAM01.1	PPB	1/19/2024 17:30	SMPL
P1063-11	E0Y02	243	0.0237 -		SFAM01.1	PPB	1/19/2024 17:32	SMPL

LB129191

INSTRUMENT ID : CV1

P1063-12	E0Y04	144	0.0044 -	SFAM01.1	PPB	1/19/2024 17:34 SMPL
P1063-13	E0WW8	276	0.0301 -	SFAM01.1	PPB	1/19/2024 17:37 SMPL
P1063-14	E0WW8D	281	0.0311 -	SFAM01.1	PPB	1/19/2024 17:39 SMPL
P1063-15	E0WW8S	5237	0.998 -	SFAM01.1	PPB	1/19/2024 17:41 SMPL
P1063-16	E0WT2	99	-0.0044 -	SFAM01.1	PPB	1/19/2024 17:43 SMPL
P1063-17	E0X08	83	-0.0075 -	SFAM01.1	PPB	1/19/2024 17:46 SMPL
CCV003	CCV003	25718	4.9936 -	SFAM01.1	PPB	1/19/2024 17:48 SMPL
CCB003	CCB003	-80	-0.0393 -	SFAM01.1	PPB	1/19/2024 17:50 SMPL
P1063-18	E0X09	86	-0.0069 -	SFAM01.1	PPB	1/19/2024 17:53 SMPL
P1063-19	E0X28	117	-0.0009 -	SFAM01.1	PPB	1/19/2024 17:55 SMPL
P1063-20	E0X29	41	-0.0157 -	SFAM01.1	PPB	1/19/2024 17:57 SMPL
P1063-21	E0Y09	57	-0.0126 -	SFAM01.1	PPB	1/19/2024 17:59 SMPL
P1063-22	E0Y10	91	-0.006 -	SFAM01.1	PPB	1/19/2024 18:02 SMPL
CCV004	CCV004	26029	5.0543 -	SFAM01.1	PPB	1/19/2024 18:04 SMPL
CCB004	CCB004	-100	-0.0432 -	SFAM01.1	PPB	1/19/2024 18:06 SMPL