



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB129194

Slope : 99.3

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	01/19/2024	16:00
2	CAL2	1	Water	NA	NA	20.3	7.00	01/19/2024	16:03
3	CAL3	1	Water	NA	NA	20.4	10.02	01/19/2024	16:05
4	ICV	1	Water	NA	NA	20.3	7.00	01/19/2024	16:10
5	CCV1	1	Water	NA	NA	20.3	2.02	01/19/2024	16:11
6	P1183-02	1	Solid	20.02	20	21.2	7.55	01/19/2024	16:15
7	P1183-02DUP	1	Solid	20.03	20	21.3	7.56	01/19/2024	16:16
8	P1183-04	1	Solid	20.02	20	21.5	7.79	01/19/2024	16:25
9	P1201-01	1	Solid	20.02	20	20.1	4.78	01/19/2024	16:30
10	P1201-02	1	Solid	20.01	20	20.2	4.85	01/19/2024	16:34
11	P1201-03	1	Solid	20.03	20	20.1	5.08	01/19/2024	16:40
12	CCV2	1	Water	NA	NA	20.2	12.02	01/19/2024	16:45

W129194

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity p1201

WorkList ID : 177277

Department : Wet-Chemistry

Date : 01-19-2024 15:40:24

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1183-02	MR-305-0-2	Solid	Corrosivity	Cool 4 deg C	HDR108	D11	01/17/2024	9045D
P1183-04	MR-305-25-27	Solid	Corrosivity	Cool 4 deg C	HDR108	D11	01/17/2024	9045D
P1201-01	3393	Solid	Corrosivity	Cool 4 deg C	PSEG03	E31	01/19/2024	9045D
P1201-02	3394	Solid	Corrosivity	Cool 4 deg C	PSEG03	E31	01/19/2024	9045D
P1201-03	3395	Solid	Corrosivity	Cool 4 deg C	PSEG03	E31	01/19/2024	9045D

Date/Time

01-19-24 15:50

Raw Sample Received by:

JD (GDC)

Raw Sample Relinquished by:

APC

Date/Time

01-19-24

Raw Sample Received by:

APC

Raw Sample Relinquished by:

APC