



Analytical Summary Report

Analysis Method: 9040C

Parameter: Corrosivity

Run Number: LB129195

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 99.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	01/19/2024	16:00
2	CAL2	1	Water	NA	NA	20.3	7.00	01/19/2024	16:03
3	CAL3	1	Water	NA	NA	20.4	10.02	01/19/2024	16:05
4	ICV	1	Water	NA	NA	20.3	7.00	01/19/2024	16:10
5	CCV1	1	Water	NA	NA	20.3	2.02	01/19/2024	16:50
6	P1201-04	1	Water	NA	NA	20.7	7.13	01/19/2024	17:00
7	P1201-04DUP	1	Water	NA	NA	20.8	7.14	01/19/2024	17:02
8	CCV2	1	Water	NA	NA	20.3	12.02	01/19/2024	17:05

129195

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity p1201 water WorkList ID : 177278 Department : Wet-Chemistry Date : 01-19-2024 15:41:35

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1201-04	3396	Water	Corrosivity	Cool 4 deg C	PSEG03	E31	01/19/2024	9040C

Date/Time 01-19-24 15:50
Raw Sample Received by: JO (CJC)
Raw Sample Relinquished by: [Signature]

Date/Time 01-19-24 18:00
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: JO (CJC)