

LB129201


7470A

INSTRUMENT ID: CV1

Linear

μ Abs.:

54493



A= 0.0000e+000

B= 1.8544e-004

C= -5.7369e-002

Rho= 0.9997534

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	
0.0	0.000	0.008	0.008	350	0.000	350					0/00
0.2	0.200	0.153	-0.047	1133	0.0 %	1133	0				-24
2.5	2.500	2.646	0.146	14578	0.0 %	14578					6
5.0	5.000	4.887	-0.113	26665	0.0 %	26665					-2
7.5	7.500	7.459	-0.041	40530	0.0 %	40530					-1
10.0	10.000	10.048	0.048	54493	0.0 %	54493					0

LB129201 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 22 Jan 2024 10:19:16

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	350	-	0.0000	7470A	PPB	22 Jan 2024 10:33:17	S	Std
0.2 - 1	50-2	1133	-	0.2000	7470A	PPB	22 Jan 2024 10:35:34	S	Std
2.5 - 1	52.5	14578	-	2.5000	7470A	PPB	22 Jan 2024 10:37:52	S	Std
5.0 - 1	55	26665	-	5.0000	7470A	PPB	22 Jan 2024 10:40:09	S	Std
7.5 - 1	57.5	40530	-	7.5000	7470A	PPB	22 Jan 2024 10:42:27	S	Std
10.0 - 1	57.5	54493	-	10.0000	7470A	PPB	22 Jan 2024 10:44:51	S	Std
ICV75 - 1	ICV75	20608	3.7642	-	7470A	PPB	22 Jan 2024 10:49:18	U	SMPL
ICB75 - 1	ICB75	243	-0.0123	-	7470A	PPB	22 Jan 2024 10:51:33	U	SMPL
CCV05 - 1	CCV05	27756	5.0897	-	7470A	PPB	22 Jan 2024 10:53:51	U	SMPL
CCB05 - 1	CCB05	199	-0.0205	-	7470A	PPB	22 Jan 2024 10:56:06	U	SMPL
CRA - 1	CRA	1129	0.1520	-	7470A	PPB	22 Jan 2024 10:58:24	U	SMPL
HighStd - 1	HighStd	54030	9.9620	-	7470A	PPB	22 Jan 2024 11:00:39	U	SMPL
ChkStd - 1	ChkStd	39521	7.2714	-	7470A	PPB	22 Jan 2024 11:02:55	U	SMPL
PB158561BL - 1	PBW	100	-0.0388	-	7470A	PPB	22 Jan 2024 11:05:15	U	SMPL
PB158561BS - 1	LCSW	24744	4.5312	-	7470A	PPB	22 Jan 2024 11:07:34	U	SMPL
P1159-03 - 1	TP-011724	328	0.0035	-	7470A	PPB	22 Jan 2024 11:09:50	U	SMPL
P1169-02 - 1	WC-A-COMP	581	0.0467	-	7470A	PPB	22 Jan 2024 11:12:09	U	SMPL
P1169-05 - 1	WC-B-COMP	433	0.0229	-	7470A	PPB	22 Jan 2024 11:14:26	U	SMPL
P1169-08 - 1	WC-C-COMP	447	0.0255	-	7470A	PPB	22 Jan 2024 11:16:43	U	SMPL
P1179-09 - 1	S-2	505	0.0363	-	7470A	PPB	22 Jan 2024 11:18:59	U	SMPL
CCV06 - 1	CCV06	26706	4.8950	-	7470A	PPB	22 Jan 2024 11:21:14	U	SMPL
CCB06 - 1	CCB06	140	-0.0314	-	7470A	PPB	22 Jan 2024 11:23:30	U	SMPL
P1179-10 - 1	S-3	386	0.0142	-	7470A	PPB	22 Jan 2024 11:25:48	U	SMPL
P1179-11 - 1	S-4	618	0.0572	-	7470A	PPB	22 Jan 2024 11:28:03	U	SMPL
P1179-12 - 1	S-5	2164	0.3439	-	7470A	PPB	22 Jan 2024 11:30:19	U	SMPL
P1179-13 - 1	S-6	415	0.0196	-	7470A	PPB	22 Jan 2024 11:32:34	U	SMPL
P1179-14 - 1	S-7	541	0.0430	-	7470A	PPB	22 Jan 2024 11:34:51	U	SMPL
P1179-15 - 1	S-8	388	0.0146	-	7470A	PPB	22 Jan 2024 11:37:07	U	SMPL
P1183-02 - 1	MR-305-0-2	344	0.0064	-	7470A	PPB	22 Jan 2024 11:39:24	U	SMPL
P1183-04 - 1	MR-305-25-27	373	0.0118	-	7470A	PPB	22 Jan 2024 11:41:41	U	SMPL
P1183-04DUP - 1	MR-305-25-27DUP	405	0.0177	-	7470A	PPB	22 Jan 2024 11:43:58	U	SMPL
P1183-04MS - 1	MR-305-25-27MS	23910	4.3765	-	7470A	PPB	22 Jan 2024 11:46:13	U	SMPL
CCV07 - 1	CCV07	27105	4.9690	-	7470A	PPB	22 Jan 2024 11:48:29	U	SMPL
CCB07 - 1	CCB07	190	-0.0221	-	7470A	PPB	22 Jan 2024 11:50:46	U	SMPL
P1183-04MSD - 1	MR-305-25-27MSD	22768	4.1647	-	7470A	PPB	22 Jan 2024 11:55:47	U	SMPL
PB158504TB - 1	PB158504TB	61	-0.0461	-	7470A	PPB	22 Jan 2024 11:58:02	U	SMPL
P1183-04LX5 - 1		2	-0.0570	-	7470A	PPB	22 Jan 2024 12:00:20	U	SMPL
P1183-04A - 1		23604	4.3198	-	7470A	PPB	22 Jan 2024 12:05:10	U	SMPL
CCV08 - 1	CCV08	27252	4.9962	-	7470A	PPB	22 Jan 2024 12:08:14	U	SMPL
CCB08 - 1	CCB08	95	-0.0398	-	7470A	PPB	22 Jan 2024 12:10:30	U	SMPL