



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 1/23/2024

OVENTEMP IN Celsius(°C): 106
Time IN: 13:25
In Date: 01/22/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:00
Out Date: 01/23/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB129202

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
P1194-01	E0X64	1	1.18	8.45	9.63	8.92	91.6	
P1194-02	E0X65	2	1.16	8.81	9.97	8.77	86.4	
P1194-03	E0X64	3	1.15	8.40	9.55	8.87	91.9	
P1194-04	E0X65	4	1.19	8.75	9.94	8.55	84.1	
P1194-05	E0X59	5	1.13	8.75	9.88	9.03	90.3	
P1194-06	E0X66	6	1.19	8.47	9.66	7.47	74.1	
P1194-07	E0X66D	7	1.19	8.47	9.66	7.47	74.1	
P1194-08	E0X66S	8	1.19	8.47	9.66	7.47	74.1	
P1194-09	E0XF1	9	1.19	8.52	9.71	8.31	83.6	
P1194-10	E0XE3	10	1.15	8.62	9.77	8.45	84.7	
P1194-11	E0XF2	11	1.19	8.63	9.82	9.03	90.8	
P1194-12	E0XF3	12	1.12	8.83	9.95	8.35	81.9	
P1194-13	E0XF4	13	1.16	8.81	9.97	6.87	64.8	
P1194-14	E0XJ7	14	1.13	8.56	9.69	7.98	80.0	
P1194-15	E0XJ9	15	1.13	8.70	9.83	8.18	81.0	
P1194-16	E0XK0	16	1.15	8.40	9.55	7.11	71.0	
P1194-17	E0Y17	17	1.13	8.80	9.93	8.92	88.5	
P1194-18	E0XD9	18	1.18	8.44	9.62	8.48	86.5	
P1194-19	E0XE0	19	1.19	8.56	9.75	9.05	91.8	
P1194-20	E0Y18	20	1.19	8.68	9.87	9.12	91.4	
P1194-21	E0XE5	21	1.18	8.53	9.71	9.02	91.9	
P1194-22	E0XF0	22	1.14	8.83	9.97	8.24	80.4	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$