



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: Rubina

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: Iwona

Run Number: LB129208

Reagent/Standard	Lot/Log #
Residual chlorine LOD, 0.05PPM	WP106191
Residual chlorine ICV-LCS, 0.4PPM	WP106190
Chlorine Calibration std, 0.1ppm	WP106185
Chlorine Calibration std, 0.2ppm	WP106186
Chlorine Calibration std, 0.8ppm	WP106187
Chlorine Calibration std, 0.0ppm	WP106193
Chlorine Calibration std, 1.6ppm	WP106188
Residual Chlorine LOQ 0.1ppm	WP106192
Residual Chlorine Calibration and CCV std, 0	WP106189
Total Chlorine Powder Pillows	W2759

Intercept: 0.0114

Slope: 1.0005

Regression: 0.999922

Seq	Lab ID	True Val (mg/l)	DF	Initial Reading	Final Reading	Difference	Result (mg/l)	%D	AnalDate	Anal Time
1	CAL1	0	1	0.000	0.000	0.000	-0.01		01/22/2024	10:00
2	CAL2	0.1	1	0.000	0.110	0.110	0.10	-1	01/22/2024	10:03
3	CAL3	0.2	1	0.000	0.220	0.220	0.21	4	01/22/2024	10:07
4	CAL4	0.4	1	0.000	0.420	0.420	0.41	2	01/22/2024	10:10
5	CAL5	0.8	1	0.000	0.810	0.810	0.80	-0.3	01/22/2024	10:14
6	CAL6	1.6	1	0.000	1.610	1.610	1.60	-0.1	01/22/2024	10:16



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: Rubina

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: Iwona

Run Number: LB129208

Seq	Lab ID	Initial Weight	Final Vol	True Value (mg/L)	DF	Initial Reading	Final Reading	Diff.	Result (mg/L)	Anal Date	Anal Time
1	ICV			0.4	1	0.0000	0.4100	0.4100	0.4000	01/22/2024	10:20
2	ICB				1	0.0000	0.0100	0.0100	0.0010	01/22/2024	10:23
3	CCV1			0.4	1	0.0000	0.4200	0.4200	0.4100	01/22/2024	10:27
4	CCB1				1	0.0000	0.0000	0.0000	-0.0090	01/22/2024	10:30
5	LB129208BL	50	50		1	0.0000	0.0000	0.0000	-0.0090	01/22/2024	10:34
6	LB129208BS	50	50	0.4	1	0.0000	0.4100	0.4100	0.4000	01/22/2024	10:38
7	P1187-07	50	50		1	0.0000	0.0700	0.0700	0.0610	01/22/2024	10:42
8	P1187-08	50	50		1	0.0000	0.1200	0.1200	0.1110	01/22/2024	10:45
9	CCV2			0.4	1	0.0000	0.4300	0.4300	0.4200	01/22/2024	10:48
10	CCB2				1	0.0000	0.0100	0.0100	0.0010	01/22/2024	10:52