

Analysis Method: 9034

ANALYST: Nikita

Parameter: Reactive Sulfide

SUPERVISOR REVIEW BY: Iwona

Run Number: LB129209

Constant: 16000

Normality1: 0.025

Normality2: 0.025

Reagent/Standard	Lot/Log #
SODIUM THIOSULFATE, 0.025N, 4LITRE	W2976
IODINE SOLUTION .025N 1L	W3050
Starch Solution, 4L	W2977

Seq	Lab ID	True Value (mg/l)	DF	Initial Weight (g)	Final Volume (ml)	T1 (ml)	T2 Initial	T2 Final	T2 Diff. (ml)	T1 - T2 Diff (mL)	Value Corrected With Blank	Result (ppm)	Anal Date	Anal Time
1	PB158536BL		1	5.00	50	2.00	0.00	1.92	1.92	0.08	0.00	0.00	01/22/2024	11:20
2	P1169-02		1	5.05	50	2.00	0.00	1.88	1.88	0.12	0.04	3.17	01/22/2024	11:23
3	P1169-02DUP		1	5.06	50	2.00	0.00	1.88	1.88	0.12	0.04	3.16	01/22/2024	11:26
4	P1169-05		1	5.03	50	2.00	0.00	1.84	1.84	0.16	0.08	6.36	01/22/2024	11:29
5	P1169-08		1	5.01	50	2.00	0.00	1.86	1.86	0.14	0.06	4.79	01/22/2024	11:32
6	P1183-02		1	5.04	50	2.00	0.00	1.90	1.90	0.10	0.02	1.59	01/22/2024	11:35
7	P1183-04		1	5.06	50	2.00	0.00	1.86	1.86	0.14	0.06	4.74	01/22/2024	11:38
8	P1201-01		1	5.04	50	2.00	0.00	1.88	1.88	0.12	0.04	3.17	01/22/2024	11:41
9	P1201-02		1	5.05	50	2.00	0.00	1.88	1.88	0.12	0.04	3.17	01/22/2024	11:44
10	P1201-03		1	5.06	50	2.00	0.00	1.82	1.82	0.18	0.10	7.91	01/22/2024	11:47

T1 = Titrant1

T2 = Titrant2

T2 Diff = T2 Final - T2 Initial

Value Corrected With Blank = ((T1 - T2 Diff) - Blank Correction(BL))

Result = ((T1 * Normality1) - ((T1 - Value Corrected With Blank) * Normality2)) * Constant / Initial Volume