



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Nikita

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB129217

Ambient Barometric Pressure (mmHg): 776.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2745

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	81.4	01/23/2024	08:30
2	P1201-01		1	11	49.00	120.2	119.2	01/23/2024	09:00
3	P1201-01DUP		1	11	49.00	120.2	119.2	01/23/2024	09:30
4	P1201-02		1	12	62.00	143.6	142.6	01/23/2024	10:00
5	P1201-03		1	2	6.00	42.8	41.8	01/23/2024	10:30

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06