



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Nikita

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB129221

Ambient Barometric Pressure (mmHg): 776.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2745

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	81.4	01/23/2024	14:40
2	P1201-04		1	13	100.00	>212.0	>212.0	01/23/2024	15:10
3	P1201-04DUP		1	12	100.00	>212.0	>212.0	01/23/2024	15:40
4	P1205-03		1	13	100.00	>212.0	>212.0	01/23/2024	16:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB-129221

WORKLIST(Hardcopy Internal Chain)

WorkList Name : flashpoint w 01/23/24 WorkList ID : 177349 Department : Wet-Chemistry Date : 01-23-2024 14:18:31

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1201-04	3396	Water	Flash Point	Cool 4 deg C	PSEG03	E31	01/19/2024	1010B
P1205-03	383	Water	Flash Point	Cool 4 deg C	PSEG03	E41	01/19/2024	1010B

Date/Time 01/23/24 14:30
Raw Sample Received by: N.P. WC
Raw Sample Relinquished by: J.D. WC

Date/Time 01/23/24 16:50
Raw Sample Received by: J.P. WC
Raw Sample Relinquished by: N.P. WC