


LB129228

7470A

INSTRUMENT ID: CV1

Linear



A= 0.0000e+000

B= 2.0473e-004

C= 2.0209e-002

Rho= 0.9995281

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.046	0.046	127	0.000	127	0				15
0.2	0.200	0.229	0.029	1021	0.0 %	1021					2
2.5	2.500	2.456	-0.044	11897	0.0 %	11897					1
5.0	5.000	5.042	0.042	24529	0.0 %	24529					3
7.5	7.500	7.279	-0.221	35456	0.0 %	35456					1
10.0	10.000	10.147	0.147	49464	0.0 %	49464					

LB129228 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 24 Jan 2024 10:44:11

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	80	127	-	0.0000	7470A	PPB	24 Jan 2024 10:54:23	S	Std
0.2 - 1	50.2	1021	-	0.2000	7470A	PPB	24 Jan 2024 10:56:40	S	Std
2.5 - 1	52.5	11897	-	2.5000	7470A	PPB	24 Jan 2024 10:58:57	S	Std
5.0 - 1	55	24529	-	5.0000	7470A	PPB	24 Jan 2024 11:01:17	S	Std
7.5 - 1	57.5	35456	-	7.5000	7470A	PPB	24 Jan 2024 11:03:35	S	Std
10.0 - 1	511	49464	-	10.0000	7470A	PPB	24 Jan 2024 11:08:33	S	Std
ICV77 - 1	ICV77	17906	3.6862	-7470A	PPB	24 Jan 2024 11:11:22	U	SMPL	
ICB77 - 1	ICB77	-204	-0.0216	-7470A	PPB	24 Jan 2024 11:13:38	U	SMPL	
CCV12 - 1	CCV12	24450	5.0260	-7470A	PPB	24 Jan 2024 11:15:56	U	SMPL	
CCB12 - 1	CCB12	-132	-0.0068	-7470A	PPB	24 Jan 2024 11:18:11	U	SMPL	
CRA - 1	CRA	955	0.2157	-7470A	PPB	24 Jan 2024 11:20:38	U	SMPL	
HighStd - 1	HighStd	48350	9.9191	-7470A	PPB	24 Jan 2024 11:22:53	U	SMPL	
ChkStd - 1	ChkStd	32234	6.6196	-7470A	PPB	24 Jan 2024 11:25:10	U	SMPL	
PB158590BL - 1	PBW	-347	-0.0508	-7470A	PPB	24 Jan 2024 11:27:30	U	SMPL	
PB158590BS - 1	LCSW	17088	3.5187	-7470A	PPB	24 Jan 2024 11:29:49	U	SMPL	
P1201-01 - 1	3393	931	0.2108	-7470A	PPB	24 Jan 2024 11:32:05	U	SMPL	
P1201-02 - 1	3394	413	0.1048	-7470A	PPB	24 Jan 2024 11:34:23	U	SMPL	
P1201-03 - 1	3395	2997	0.6338	-7470A	PPB	24 Jan 2024 11:36:44	U	SMPL	
PB158558TB - 1	PB158558TB	-16	0.0169	-7470A	PPB	24 Jan 2024 11:48:20	U	SMPL	
CCV13 - 1	CCV13	24194	4.9736	-7470A	PPB	24 Jan 2024 11:50:35	U	SMPL	
CCB13 - 1	CCB13	-294	-0.0400	-7470A	PPB	24 Jan 2024 11:52:50	U	SMPL	