



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 1/25/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 17:10
In Date: 01/24/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:15
Out Date: 01/25/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB129232

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
P1237-01	SB-2-9.5-10.0	1	1.15	8.84	9.99	8.79	86.4	
P1237-02	SB-2-19.5-20.0	2	1.13	8.50	9.63	8.16	82.7	
P1240-01	72-11985	12	1.15	8.63	9.78	8.9	89.8	
P1240-02	72-11985-EPH-2	13	1.13	8.60	9.73	8.85	89.8	
P1240-04	VNJ-232	14	1.17	8.57	9.74	8.49	85.4	
P1240-05	VNJ-232-EPH-2	15	1.19	8.43	9.62	8.27	84.0	
P1240-07	VNJ-223	16	1.19	8.56	9.75	8.75	88.3	
P1240-08	VNJ-223-EPH-2	17	1.12	8.46	9.58	8.63	88.8	
P1241-01	AU-05-12424	18	1.13	8.65	9.78	8.4	84.0	
P1242-01	TR-04-012424	19	1.15	8.81	9.96	8.64	85.0	
P1242-02	TR-04-012424	20	1.16	8.45	9.61	8.41	85.8	
P1243-01	MR-308-0-2	21	1.15	8.81	9.96	9.12	90.5	
P1243-03	MR-308-17-19	22	1.11	8.48	9.59	8.89	91.7	
P1244-01	69-TP-3	3	1.19	8.53	9.72	8.57	86.5	
P1244-02	69-TP-3-E2	4	1.16	8.78	9.94	8.66	85.4	
P1245-01	MIXED-DEMO	5	1.15	8.37	9.52	9.36	98.1	
P1246-01	CONCRETE-012424	6	1.00	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
P1246-02	TP-COMP-012424	7	1.16	8.67	9.83	9.03	90.8	
P1247-01	CL-02-012424	8	1.13	8.61	9.74	8.6	86.8	
P1247-02	CL-02-012424-E2	9	1.19	8.60	9.79	8.74	87.8	
P1248-01	EO-02-012424	10	1.11	8.67	9.78	8.73	87.9	
P1248-02	EO-02-012424-E2	11	1.13	8.75	9.88	8.74	87.0	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$