

LB129233


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

61649



A= 0.0000e+000

B= 1.6332e-004

C= 1.2421e-002

Rho= 0.9998784

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%SD
0.0	0.000	0.041	0.041	174	0.000	174					
0.2	0.200	0.200	0.000	1149	0.0 %	1149					
2.5	2.500	2.482	-0.018	15120	0.0 %	15120					
5.0	5.000	5.003	0.003	30554	0.0 %	30554					
7.5	7.500	7.394	-0.106	45195	0.0 %	45195					
10.0	10.000	10.081	0.081	61649	0.0 %	61649					

10.011

LB129233 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 24 Jan 2024 12:49:25

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	174	-	0.0000	7471B	PPB	24 Jan 2024 13:13:08	S	Std
0.2 - 1	50.2	1149	-	0.2000	7471B	PPB	24 Jan 2024 13:15:25	S	Std
2.5 - 1	52.5	15120	-	2.5000	7471B	PPB	24 Jan 2024 13:17:42	S	Std
5.0 - 1	55	30554	-	5.0000	7471B	PPB	24 Jan 2024 13:19:59	S	Std
7.5 - 1	57.5	45195	-	7.5000	7471B	PPB	24 Jan 2024 13:25:04	S	Std
10.0 - 1	514	61649	-	10.0000	7471B	PPB	24 Jan 2024 13:27:23	S	Std
ICV78 - 1	ICV78	24333	3.9865	-	7471B	PPB	24 Jan 2024 13:30:18	U	SMPL
ICB78 - 1	ICB78	-66	0.0016	-	7471B	PPB	24 Jan 2024 13:32:35	U	SMPL
CCV14 - 1	CCV14	30800	5.0427	-	7471B	PPB	24 Jan 2024 13:34:58	U	SMPL
CCB14 - 1	CCB14	-82	-0.0010	-	7471B	PPB	24 Jan 2024 13:37:13	U	SMPL
CRA - 1	CRA	1126	0.1963	-	7471B	PPB	24 Jan 2024 13:39:39	U	SMPL
HighStd - 1	HighStd	62264	10.1814	-	7471B	PPB	24 Jan 2024 13:41:54	U	SMPL
ChkStd - 1	ChkStd	42674	6.9820	-	7471B	PPB	24 Jan 2024 13:44:09	U	SMPL
PB158609BL - 1	PBS	-115	-0.0064	-	7471B	PPB	24 Jan 2024 13:46:40	U	SMPL
PB158609BS - 1	LCSS	22188	3.6362	-	7471B	PPB	24 Jan 2024 13:51:50	U	SMPL
P1218-03 - 1	1029	3885	0.6469	-	7471B	PPB	24 Jan 2024 13:54:07	U	SMPL
P1218-04 - 1	133	1264	0.2189	-	7471B	PPB	24 Jan 2024 13:56:30	U	SMPL
P1218-05 - 1	134	2309	0.3895	-	7471B	PPB	24 Jan 2024 13:58:49	U	SMPL
P1219-01 - 1	222177	2235	0.3774	-	7471B	PPB	24 Jan 2024 14:01:06	U	SMPL
P1221-01 - 1	MH-B	549	0.1021	-	7471B	PPB	24 Jan 2024 14:03:24	U	SMPL
CCV15 - 1	CCV15	30083	4.9256	-	7471B	PPB	24 Jan 2024 14:05:41	U	SMPL
CCB15 - 1	CCB15	-77	-0.0002	-	7471B	PPB	24 Jan 2024 14:07:56	U	SMPL
P1221-03 - 1	MH-A	1395	0.2403	-	7471B	PPB	24 Jan 2024 14:10:20	U	SMPL
P1221-03DUP - 1	MH-ADUP	1357	0.2340	-	7471B	PPB	24 Jan 2024 14:12:36	U	SMPL
P1221-03MS - 1	MH-AMS	25234	4.1337	-	7471B	PPB	24 Jan 2024 14:14:52	U	SMPL
P1221-03MSD - 1	MH-AMSD	24747	4.0541	-	7471B	PPB	24 Jan 2024 14:17:07	U	SMPL
P1221-03LX5 - 1		195	0.0443	-	7471B	PPB	24 Jan 2024 14:22:10	U	SMPL
P1221-03A - 1		24619	4.0332	-	7471B	PPB	24 Jan 2024 14:24:27	U	SMPL
CCV16 - 1	CCV16	30713	5.0285	-	7471B	PPB	24 Jan 2024 14:26:43	U	SMPL
CCB16 - 1	CCB16	-63	0.0021	-	7471B	PPB	24 Jan 2024 14:29:06	U	SMPL