



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: Rubina

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: Iwona

Run Number: LB129237

Reagent/Standard	Lot/Log #
Residual chlorine LOD, 0.05PPM	WP106265
Residual chlorine ICV-LCS, 0.4PPM	WP106264
Chlorine Calibration std, 0.1ppm	WP106259
Chlorine Calibration std, 0.2ppm	WP106260
Chlorine Calibration std, 0.8ppm	WP106261
Chlorine Calibration std, 0.0ppm	WP106258
Chlorine Calibration std, 1.6ppm	WP106262
Residual Chlorine Calibration and CCV std, 0	WP106263
Total Chlorine Powder Pillows	W2759

Intercept: 0.0120

Slope: 1.0090

Regression: 0.999841

Seq	Lab ID	True Val (mg/l)	DF	Initial Reading	Final Reading	Difference	Result (mg/l)	%D	AnalDate	Anal Time
1	CAL1	0	1	0.000	0.000	0.000	-0.01		01/24/2024	12:30
2	CAL2	0.1	1	0.000	0.120	0.120	0.11	7	01/24/2024	12:34
3	CAL3	0.2	1	0.000	0.230	0.230	0.22	8	01/24/2024	12:38
4	CAL4	0.4	1	0.000	0.410	0.410	0.39	-1.5	01/24/2024	12:42
5	CAL5	0.8	1	0.000	0.810	0.810	0.79	-1.1	01/24/2024	12:45
6	CAL6	1.6	1	0.000	1.630	1.630	1.60	0.3	01/24/2024	12:48



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: Rubina

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: Iwona

Run Number: LB129237

Seq	Lab ID	Initial Weight	Final Vol	True Value (mg/L)	DF	Initial Reading	Final Reading	Diff.	Result (mg/L)	Anal Date	Anal Time
1	ICV			0.4	1	0.0000	0.4300	0.4300	0.4140	01/24/2024	12:52
2	ICB				1	0.0000	0.0200	0.0200	0.0080	01/24/2024	12:56
3	CCV1			0.4	1	0.0000	0.4200	0.4200	0.4040	01/24/2024	13:00
4	CCB1				1	0.0000	0.0100	0.0100	-0.0020	01/24/2024	13:03
5	LB129237BL	50	50		1	0.0000	0.0000	0.0000	-0.0120	01/24/2024	13:07
6	LB129237BS	50	50	0.4	1	0.0000	0.4100	0.4100	0.3940	01/24/2024	13:10
7	P1187-09	50	50		1	0.0000	0.0600	0.0600	0.0480	01/24/2024	13:14
8	CCV2			0.4	1	0.0000	0.4200	0.4200	0.4040	01/24/2024	13:18
9	CCB2				1	0.0000	0.0000	0.0000	-0.0120	01/24/2024	13:22