

SFAM01.1


LB129241

INSTRUMENT ID: CV1

Linear

μ Abs.:

52322



A= 0.0000e+000

B= 1.9307e-004 slope

C= -5.2293e-002 tolerance

Rho= 0.9996949

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	c/d
0.00	0.000	0.012	0.012	333	0.000	333					0
0.05	0.050					442	0				1
0.20	0.200	0.200	0.000	1309	0.0 %	1309					2
2.50	2.500	2.443	-0.057	12925	0.0 %	12925					3
5.00	5.000	5.146	0.146	26924	0.0 %	26924					2
7.50	7.500	7.349	-0.151	38336	0.0 %	38336					0
10.0	10.000	10.049	0.049	52322	0.0 %	52322					

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	333	-		0 SFAM01.1	PPB	1/24/2024 15:51	Std
	0.2 S01	1309	-		0.2 SFAM01.1	PPB	1/24/2024 15:53	Std
	2.5 S02	12925	-		2.5 SFAM01.1	PPB	1/24/2024 15:55	Std
	5 S03	26924	-		5 SFAM01.1	PPB	1/24/2024 15:57	Std
	7.5 S04	38336	-		7.5 SFAM01.1	PPB	1/24/2024 16:00	Std
	10 S05	52322	-		10 SFAM01.1	PPB	1/24/2024 16:02	Std
ICV079	ICV079	20410	3.8882 -		SFAM01.1	PPB	1/24/2024 16:05	SMPL
ICB079	ICB079	196	-0.0145 -		SFAM01.1	PPB	1/24/2024 16:07	SMPL
CCV017	CCV017	27111	5.182 -		SFAM01.1	PPB	1/24/2024 16:09	SMPL
CCB017	CCB017	191	-0.0154 -		SFAM01.1	PPB	1/24/2024 16:12	SMPL
PB158592BL	PBW592	594	0.0624 -		SFAM01.1	PPB	1/24/2024 16:14	SMPL
P1092-01	MCOAW8	337	0.0128 -		SFAM01.1	PPB	1/24/2024 16:16	SMPL
P1092-02	MCOAX6	274	0.0006 -		SFAM01.1	PPB	1/24/2024 16:19	SMPL
P1092-03	MCOAX7	358	0.0168 -		SFAM01.1	PPB	1/24/2024 16:21	SMPL
P1092-04	MCOAZ1	342	0.0137 -		SFAM01.1	PPB	1/24/2024 16:23	SMPL
P1092-05	MCOB55	370	0.0191 -		SFAM01.1	PPB	1/24/2024 16:25	SMPL
P1092-06	MCOB55D	342	0.0137 -		SFAM01.1	PPB	1/24/2024 16:28	SMPL
P1092-07	MCOB55S	4994	0.9119 -		SFAM01.1	PPB	1/24/2024 16:30	SMPL
P1092-08	MCOAX5	278	0.0014 -		SFAM01.1	PPB	1/24/2024 16:32	SMPL
P1092-09	MCOAZ3	297	0.005 -		SFAM01.1	PPB	1/24/2024 16:34	SMPL
P1092-10	MCOAZ7	341	0.0135 -		SFAM01.1	PPB	1/24/2024 16:37	SMPL
P1092-11	MCOB01	1719	0.2796 -		SFAM01.1	PPB	1/24/2024 16:39	SMPL
P1092-12	MCOB37	1408	0.2195 -		SFAM01.1	PPB	1/24/2024 16:41	SMPL
P1092-13	MCOB57	1537	0.2445 -		SFAM01.1	PPB	1/24/2024 16:43	SMPL
P1092-14	MCOAY4	1179	0.1753 -		SFAM01.1	PPB	1/24/2024 16:46	SMPL
P1092-15	MCOAZ6	320	0.0095 -		SFAM01.1	PPB	1/24/2024 16:48	SMPL
P1092-16	MCOB38	1825	0.3001 -		SFAM01.1	PPB	1/24/2024 16:50	SMPL
P1092-17	MCOB59	2217	0.3757 -		SFAM01.1	PPB	1/24/2024 16:53	SMPL
P1092-18	MCOB04	2649	0.4591 -		SFAM01.1	PPB	1/24/2024 16:55	SMPL
PB158593BL	PBW593	573	0.0583 -		SFAM01.1	PPB	1/24/2024 16:57	SMPL
CCV018	CCV018	26914	5.1439 -		SFAM01.1	PPB	1/24/2024 16:59	SMPL
CCB018	CCB018	176	-0.0183 -		SFAM01.1	PPB	1/24/2024 17:02	SMPL
P1093-01	MCOB15	75	-0.0378 -		SFAM01.1	PPB	1/24/2024 17:04	SMPL
P1093-02	MCOB16	107	-0.0316 -		SFAM01.1	PPB	1/24/2024 17:06	SMPL
P1093-03	MCOB17	125	-0.0282 -		SFAM01.1	PPB	1/24/2024 17:08	SMPL
P1093-04	MCOB30	126	-0.028 -		SFAM01.1	PPB	1/24/2024 17:11	SMPL
P1093-05	MCOB31	150	-0.0233 -		SFAM01.1	PPB	1/24/2024 17:13	SMPL
P1093-06	MCOB56	158	-0.0218 -		SFAM01.1	PPB	1/24/2024 17:15	SMPL
P1093-07	MCOB56D	155	-0.0224 -		SFAM01.1	PPB	1/24/2024 17:17	SMPL
P1093-08	MCOB56S	6176	1.1401 -		SFAM01.1	PPB	1/24/2024 17:20	SMPL
P1093-09	MCOB18	97	-0.0336 -		SFAM01.1	PPB	1/24/2024 17:22	SMPL
P1093-10	MCOB42	154	-0.0226 -		SFAM01.1	PPB	1/24/2024 17:24	SMPL
P1093-11	MCOB58	160	-0.0214 -		SFAM01.1	PPB	1/24/2024 17:27	SMPL
P1093-12	MCOB24	202	-0.0133 -		SFAM01.1	PPB	1/24/2024 17:29	SMPL
P1093-13	MCOB39	171	-0.0193 -		SFAM01.1	PPB	1/24/2024 17:31	SMPL
P1093-14	MCOB60	229	-0.0081 -		SFAM01.1	PPB	1/24/2024 17:33	SMPL

LB129241

INSTRUMENT ID : CV1

P1093-15	MCOB61	276	0.001 -	SFAM01.1	PPB	1/24/2024 17:36 SMPL
P1093-17	MCOB32	457	0.0359 -	SFAM01.1	PPB	1/24/2024 17:38 SMPL
P1093-18	MCOB33	1508	0.2389 -	SFAM01.1	PPB	1/24/2024 17:40 SMPL
PB158612BL	PBW612	202	-0.0133 -	SFAM01.1	PPB	1/24/2024 17:42 SMPL
P1114-01	MCOB45	338	0.013 -	SFAM01.1	PPB	1/24/2024 17:45 SMPL
P1114-02	MCOB45D	1146	0.169 -	SFAM01.1	PPB	1/24/2024 17:47 SMPL
CCV019	CCV019	26396	5.0439 -	SFAM01.1	PPB	1/24/2024 17:49 SMPL
CCB019	CCB019	45	-0.0436 -	SFAM01.1	PPB	1/24/2024 17:52 SMPL
P1114-03	MCOB45S	7005	1.3001 -	SFAM01.1	PPB	1/24/2024 17:54 SMPL
P1114-05	MCOAZ0	1240	0.1871 -	SFAM01.1	PPB	1/24/2024 17:56 SMPL
P1114-06	MCOAZ2	2909	0.5093 -	SFAM01.1	PPB	1/24/2024 17:58 SMPL
P1114-07	MCOB64	1685	0.273 -	SFAM01.1	PPB	1/24/2024 18:01 SMPL
P1114-08	MCOB66	1696	0.2752 -	SFAM01.1	PPB	1/24/2024 18:03 SMPL
P1115-01	MCOB46	1575	0.2518 -	SFAM01.1	PPB	1/24/2024 18:05 SMPL
P1115-02	MCOB46D	1616	0.2597 -	SFAM01.1	PPB	1/24/2024 18:07 SMPL
P1115-03	MCOB46S	6764	1.2536 -	SFAM01.1	PPB	1/24/2024 18:10 SMPL
P1127-01	EOY13	66	-0.0396 -	SFAM01.1	PPB	1/24/2024 18:12 SMPL
CCV020	CCV020	26523	5.0685 -	SFAM01.1	PPB	1/24/2024 18:14 SMPL
CCB020	CCB020	46	-0.0434 -	SFAM01.1	PPB	1/24/2024 18:17 SMPL