



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Nikita

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB129247

Ambient Barometric Pressure (mmHg): 776.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2745

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	81.4	01/26/2024	08:20
2	P1218-07		1	11	60.00	140.0	139.0	01/26/2024	08:50
3	P1218-08		1	4	7.00	44.6	43.6	01/26/2024	09:20
4	P1226-01		1	13	100.00	>212.0	>212.0	01/26/2024	09:50
5	P1226-03		1	12	100.00	>212.0	>212.0	01/26/2024	10:20
6	P1235-09		1	11	52.00	125.6	124.6	01/26/2024	10:50
7	P1236-01		1	13	100.00	>212.0	>212.0	01/26/2024	11:20
8	P1236-01DUP		1	12	100.00	>212.0	>212.0	01/26/2024	11:50

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB-129247

WORKLIST(Hardcopy Internal Chain)

WorkList Name : 01/24/24 flashpoint WorkList ID : 177374 Department : Wet-Chemistry Date : 01-24-2024 14:56:04

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1218-07	132	Solid	Flash Point	Cool 4 deg C	PSEG03	E41	01/22/2024	1010B
P1218-08	1155	Solid	Flash Point	Cool 4 deg C	PSEG03	E41	01/22/2024	1010B
P1226-01	1219	Solid	Flash Point	Cool 4 deg C	PSEG03	E41	01/23/2024	1010B
P1226-03	0919	Solid	Flash Point	Cool 4 deg C	PSEG03	E41	01/23/2024	1010B
P1235-09	136	Solid	Flash Point	Cool 4 deg C	PSEG03	E41	01/23/2024	1010B
P1236-01	1204	Solid	Flash Point	Cool 4 deg C	PSEG03	E42	01/23/2024	1010B

Date/Time 01/26/2024 8:10
Raw Sample Received by: N.P. W.C.
Raw Sample Relinquished by: [Signature]

Date/Time 01/26/2024 12:30
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: N.P. W.C.