

LB129253


SFAM01.1

INSTRUMENT 1A: CVJ

Linear

 μ Abs.:

55363



A= 0.0000e+000

B= 1.7956e-004 slope

C= 3.9584e-002 *intercept*

Rho= 0.9998960

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	
0.00	0.000	0.062	0.062	125	0.000	125					
0.05	0.050					442	0				
0.20	0.200	0.198	-0.002	884	0.0 %	884					
2.50	2.500	2.420	-0.080	13257	0.0 %	13257					
5.00	5.000	4.966	-0.034	27437	0.0 %	27437					
7.50	7.500	7.572	0.072	41951	0.0 %	41951					
10.0	10.000	9.981	-0.019	55363	0.0 %	55363					

0/00
1
1
3
1
1
0

LB129253

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	125	-		0 SFAM01.1	PPB	1/25/2024 14:11	Std
	0.2 S01	884	-		0.2 SFAM01.1	PPB	1/25/2024 14:14	Std
	2.5 S02	13257	-		2.5 SFAM01.1	PPB	1/25/2024 14:16	Std
	5 S03	27437	-		5 SFAM01.1	PPB	1/25/2024 14:18	Std
	7.5 S04	41951	-		7.5 SFAM01.1	PPB	1/25/2024 14:20	Std
	10 S05	55363	-		10 SFAM01.1	PPB	1/25/2024 14:23	Std
ICV080	ICV080	20714	3.7591 -		SFAM01.1	PPB	1/25/2024 14:26	SMPL
ICB080	ICB080	-95	0.0225 -		SFAM01.1	PPB	1/25/2024 14:28	SMPL
CCV021	CCV021	27290	4.9399 -		SFAM01.1	PPB	1/25/2024 14:30	SMPL
CCB021	CCB021	-107	0.0204 -		SFAM01.1	PPB	1/25/2024 14:32	SMPL
PB158639BL	PBW639	82	0.0543 -		SFAM01.1	PPB	1/25/2024 14:35	SMPL
O5867-10	GCEW3	31383	5.6749 -		SFAM01.1	PPB	1/25/2024 14:37	SMPL
PB158640BL	PBW640	-117	0.0186 -		SFAM01.1	PPB	1/25/2024 14:39	SMPL
P1112-01	MCOAY9	20	0.0432 -		SFAM01.1	PPB	1/25/2024 14:42	SMPL
P1112-02	MCOAY9D	58	0.05 -		SFAM01.1	PPB	1/25/2024 14:44	SMPL
P1112-03	MCOAY9S	5542	1.0347 -		SFAM01.1	PPB	1/25/2024 14:46	SMPL
P1112-04	MCOAY2	-81	0.025 -		SFAM01.1	PPB	1/25/2024 14:48	SMPL
P1112-05	MCOAY3	12	0.0417 -		SFAM01.1	PPB	1/25/2024 14:51	SMPL
P1112-06	MCOAY5	945	0.2093 -		SFAM01.1	PPB	1/25/2024 14:53	SMPL
P1112-07	MCOAY6	-99	0.0218 -		SFAM01.1	PPB	1/25/2024 14:55	SMPL
P1112-08	MCOAY7	291	0.0918 -		SFAM01.1	PPB	1/25/2024 14:57	SMPL
P1112-09	MCOAY8	198	0.0751 -		SFAM01.1	PPB	1/25/2024 15:00	SMPL
P1112-10	MCOAZ8	41	0.0469 -		SFAM01.1	PPB	1/25/2024 15:02	SMPL
P1112-11	MCOAZ9	69	0.052 -		SFAM01.1	PPB	1/25/2024 15:04	SMPL
P1112-12	MCOB03	67	0.0516 -		SFAM01.1	PPB	1/25/2024 15:06	SMPL
P1112-13	MCOAZ4	99	0.0574 -		SFAM01.1	PPB	1/25/2024 15:09	SMPL
P1112-14	MCOB02	101	0.0577 -		SFAM01.1	PPB	1/25/2024 15:11	SMPL
P1112-15	MCOAW9	323	0.0976 -		SFAM01.1	PPB	1/25/2024 15:13	SMPL
P1112-16	MCOAX0	157	0.0678 -		SFAM01.1	PPB	1/25/2024 15:16	SMPL
P1112-17	MCOAX2	72	0.0525 -		SFAM01.1	PPB	1/25/2024 15:18	SMPL
CCV022	CCV022	25980	4.7047 -		SFAM01.1	PPB	1/25/2024 15:20	SMPL
CCB022	CCB022	-128	0.0166 -		SFAM01.1	PPB	1/25/2024 15:22	SMPL
P1112-19	MCOAY0	983	0.2161 -		SFAM01.1	PPB	1/25/2024 15:25	SMPL
P1112-20	MCOAY1	65	0.0513 -		SFAM01.1	PPB	1/25/2024 15:27	SMPL
P1112-22	MCOB00	75	0.0531 -		SFAM01.1	PPB	1/25/2024 15:29	SMPL
PB158641BL	PBW641	126	0.0622 -		SFAM01.1	PPB	1/25/2024 15:31	SMPL
P1113-01	MCOB29	114	0.0601 -		SFAM01.1	PPB	1/25/2024 15:34	SMPL
P1113-02	MCOB29D	140	0.0647 -		SFAM01.1	PPB	1/25/2024 15:36	SMPL
P1113-03	MCOB29S	4944	0.9273 -		SFAM01.1	PPB	1/25/2024 15:38	SMPL
P1113-04	MCOB22	29	0.0448 -		SFAM01.1	PPB	1/25/2024 15:40	SMPL
P1113-05	MCOB23	110	0.0593 -		SFAM01.1	PPB	1/25/2024 15:43	SMPL
P1113-06	MCOB28	179	0.0717 -		SFAM01.1	PPB	1/25/2024 15:45	SMPL
P1113-07	MCOB26	215	0.0782 -		SFAM01.1	PPB	1/25/2024 15:47	SMPL
P1113-08	MCOB27	194	0.0744 -		SFAM01.1	PPB	1/25/2024 15:50	SMPL
P1113-09	MCOB09	112	0.0597 -		SFAM01.1	PPB	1/25/2024 15:52	SMPL
P1113-10	MCOB10	109	0.0592 -		SFAM01.1	PPB	1/25/2024 15:54	SMPL

LB129253

INSTRUMENT ID : CV1

P1113-11	MC0B12	94	0.0565 -	SFAM01.1	PPB	1/25/2024 15:56 SMPL
P1113-12	MC0B20	432	0.1172 -	SFAM01.1	PPB	1/25/2024 15:59 SMPL
P1113-13	MC0B21	104	0.0583 -	SFAM01.1	PPB	1/25/2024 16:01 SMPL
P1113-14	MC0B25	139	0.0645 -	SFAM01.1	PPB	1/25/2024 16:03 SMPL
P1113-15	MC0B44	115	0.0602 -	SFAM01.1	PPB	1/25/2024 16:05 SMPL
CCV023	CCV023	25338	4.5894 -	SFAM01.1	PPB	1/25/2024 16:12 SMPL
CCB023	CCB023	-200	0.0037 -	SFAM01.1	PPB	1/25/2024 16:15 SMPL