



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB129257

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	01/25/2024	15:10
2	CAL2	1	Water	NA	NA	20.3	7.00	01/25/2024	15:11
3	CAL3	1	Water	NA	NA	20.4	10.02	01/25/2024	15:15
4	ICV	1	Water	NA	NA	20.4	7.02	01/25/2024	15:20
5	CCV1	1	Water	NA	NA	20.3	2.01	01/25/2024	15:25
6	P1256-01	1	Water	NA	NA	20.7	6.69	01/25/2024	15:30
7	P1258-01	1	Water	NA	NA	20.3	6.15	01/25/2024	15:35
8	P1258-02	1	Water	NA	NA	20.3	5.79	01/25/2024	15:40
9	P1258-03	1	Water	NA	NA	20.3	5.98	01/25/2024	15:45
10	P1258-03DUP	1	Water	NA	NA	20.4	5.99	01/25/2024	15:46
11	CCV2	1	Water	NA	NA	20.4	12.02	01/25/2024	15:50

WORKLIST(Hardcopy Internal Chain)

VB 129257

WorkList Name : PH P1256 WATER

WorkList ID : 177405

Department : Wet-Chemistry

Date : 01-25-2024 14:51:30

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1256-01 A	60131	Water	pH	Cool 4 deg C	PSEG03	E51	01/25/2024	9040C
P1258-01 C	50234	Water	pH	Cool 4 deg C	PSEG03	E41	01/25/2024	9040C
P1258-02 C	50235	Water	pH	Cool 4 deg C	PSEG03	E41	01/25/2024	9040C
P1258-03 C	50236	Water	pH	Cool 4 deg C	PSEG03	E41	01/25/2024	9040C

Date/Time 01-25-24 15:00

Raw Sample Received by: JDC

Raw Sample Relinquished by: JDC

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

VB 01-25-24 17:30
JDC JDC