



### Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB129258

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	01/25/2024	15:10
2	CAL2	1	Water	NA	NA	20.3	7.00	01/25/2024	15:11
3	CAL3	1	Water	NA	NA	20.4	10.02	01/25/2024	15:15
4	ICV	1	Water	NA	NA	20.4	7.02	01/25/2024	15:20
5	CCV1	1	Water	NA	NA	20.3	2.01	01/25/2024	16:00
6	P1241-01	1	Solid	20.02	20	23.5	8.03	01/25/2024	16:10
7	P1241-01DUP	1	Solid	20.03	20	23.6	8.04	01/25/2024	16:11
8	P1257-01	1	Solid	20.05	20	24.2	6.73	01/25/2024	16:15
9	CCV2	1	Water	NA	NA	20.3	12.02	01/25/2024	16:25

UP 129258

### WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH P1256 SOLIDS      WorkList ID : 177406      Department : Wet-Chemistry      Date : 01-25-2024 14:52:40

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1241-01	AU-05-12424	Solid	pH	Cool 4 deg C	PSEG05	E41	01/24/2024	9045D
P1257-01	34626	Solid	pH	Cool 4 deg C	PSEG03	E51	01/25/2024	9045D

Date/Time 01/25/24 15:00  
 Raw Sample Received by: JS (coc)  
 Raw Sample Relinquished by: JS

Date/Time 01/25/24      17:30  
 Raw Sample Received by: JS  
 Raw Sample Relinquished by: JS (coc)