

LB129260


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

56573



A= 0.0000e+000

B= 1.7977e-004

C= 3.9882e-002

Rho= 0.9989319

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.065	0.065	141	0.000	141					-
0.2	0.200	0.196	-0.004	867	0.0 %	867					-2
2.5	2.500	2.518	0.018	13787	0.0 %	13787					1
5.0	5.000	5.060	0.060	27924	0.0 %	27924					1
7.5	7.500	7.151	-0.349	39554	0.0 %	39554					-5
10.0	10.000	10.210	0.210	56573	0.0 %	56573					2

LB129260

Method: 7471B

Operator: Admin

Date of Analysis: 26 Jan 2024 10:04:17

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	141	-	0.00007471B	PPB		26 Jan 2024 10:16:34	S	Std
0.2 - 1	50.2	867	-	0.20007471B	PPB		26 Jan 2024 10:18:51	S	Std
2.5 - 1	52.5	13787	-	2.50007471B	PPB		26 Jan 2024 10:21:09	S	Std
5.0 - 1	55	27924	-	5.00007471B	PPB		26 Jan 2024 10:23:26	S	Std
7.5 - 1	57.5	39554	-	7.50007471B	PPB		26 Jan 2024 10:25:46	S	Std
10.0 - 1	111	56573	-	10.00007471B	PPB		26 Jan 2024 10:36:47	S	Std
ICV81 - 1	ICV81	23596	4.2818	-7471B	PPB		26 Jan 2024 10:39:36	U	SMPL
ICB81 - 1	ICB81	-293	-0.0128	-7471B	PPB		26 Jan 2024 10:41:53	U	SMPL
CCV24 - 1	CCV24	28773	5.2125	-7471B	PPB		26 Jan 2024 10:44:16	U	SMPL
CCB24 - 1	CCB24	-278	-0.0101	-7471B	PPB		26 Jan 2024 10:46:32	U	SMPL
CRA - 1	CRA	905	0.2026	-7471B	PPB		26 Jan 2024 10:48:58	U	SMPL
HighStd - 1	HighStd	57130	10.3103	-7471B	PPB		26 Jan 2024 10:51:13	U	SMPL
ChkStd - 1	ChkStd	36965	6.6852	-7471B	PPB		26 Jan 2024 10:53:30	U	SMPL
PB158645BL - 1	PBS	-340	-0.0212	-7471B	PPB		26 Jan 2024 10:55:58	U	SMPL
PB158645BS - 1	LCSS	21539	3.9120	-7471B	PPB		26 Jan 2024 11:00:58	U	SMPL
P1233-01 - 1	69-TP-1	3901	0.7412	-7471B	PPB		26 Jan 2024 11:03:17	U	SMPL
P1234-01 - 1	293	4927	0.9256	-7471B	PPB		26 Jan 2024 11:05:40	U	SMPL
P1235-01 - 1	138-141	4086	0.7744	-7471B	PPB		26 Jan 2024 11:07:58	U	SMPL
P1235-03 - 1	142-144	1945	0.3895	-7471B	PPB		26 Jan 2024 11:10:18	U	SMPL
P1235-05 - 1	1104	245	0.0839	-7471B	PPB		26 Jan 2024 11:12:35	U	SMPL
CCV25 - 1	CCV25	29253	5.2988	-7471B	PPB		26 Jan 2024 11:14:51	U	SMPL
CCB25 - 1	CCB25	-162	0.0108	-7471B	PPB		26 Jan 2024 11:17:07	U	SMPL
P1235-06 - 1	137	975	0.2152	-7471B	PPB		26 Jan 2024 11:19:32	U	SMPL
P1235-07 - 1	131	350	0.1028	-7471B	PPB		26 Jan 2024 11:21:48	U	SMPL
P1235-08 - 1	1101	333	0.0997	-7471B	PPB		26 Jan 2024 11:24:08	U	SMPL
P1236-01 - 1	1204	19629	3.5686	-7471B	PPB		26 Jan 2024 11:26:29	U	SMPL
P1240-01 - 1	72-11985	2869	0.5556	-7471B	PPB		26 Jan 2024 11:28:44	U	SMPL
P1240-04 - 1	VNJ-232	807	0.1850	-7471B	PPB		26 Jan 2024 11:31:08	U	SMPL
P1240-07 - 1	VNJ-223	1560	0.3203	-7471B	PPB		26 Jan 2024 11:33:26	U	SMPL
P1241-01 - 1	AU-05-12424	1906	0.3825	-7471B	PPB		26 Jan 2024 11:35:43	U	SMPL
P1242-01 - 1	TR-04-012424	2687	0.5229	-7471B	PPB		26 Jan 2024 11:38:01	U	SMPL
P1243-01 - 1	MR-308-0-2	1417	0.2946	-7471B	PPB		26 Jan 2024 11:40:17	U	SMPL
CCV26 - 1	CCV26	29458	5.3356	-7471B	PPB		26 Jan 2024 11:42:34	U	SMPL
CCB26 - 1	CCB26	-137	0.0153	-7471B	PPB		26 Jan 2024 11:44:49	U	SMPL
P1243-03 - 1	MR-308-17-19	3028	0.5842	-7471B	PPB		26 Jan 2024 11:47:14	U	SMPL
P1244-01 - 1	69-TP-3	561	0.1407	-7471B	PPB		26 Jan 2024 11:49:29	U	SMPL
P1246-02 - 1	TP-COMP-012424	790	0.1819	-7471B	PPB		26 Jan 2024 11:51:46	U	SMPL
P1246-02DUP - 1	TP-COMP-012424DUP	643	0.1555	-7471B	PPB		26 Jan 2024 11:54:02	U	SMPL
P1246-02MS - 1	TP-COMP-012424MS	22535	4.0911	-7471B	PPB		26 Jan 2024 11:56:18	U	SMPL
P1246-02MSD - 1	TP-COMP-012424MSD	22322	4.0528	-7471B	PPB		26 Jan 2024 12:01:50	U	SMPL
P1247-01 - 1	CL-02-012424	6649	1.2352	-7471B	PPB		26 Jan 2024 12:04:06	U	SMPL
P1248-01 - 1	EO-02-012424	1742	0.3530	-7471B	PPB		26 Jan 2024 12:06:26	U	SMPL
P1246-02LX5 - 1		134	0.0640	-7471B	PPB		26 Jan 2024 12:08:45	U	SMPL
P1246-02A - 1		24267	4.4024	-7471B	PPB		26 Jan 2024 12:11:01	U	SMPL
CCV27 - 1	CCV27	29191	5.2876	-7471B	PPB		26 Jan 2024 12:13:16	U	SMPL
CCB27 - 1	CCB27	-120	0.0183	-7471B	PPB		26 Jan 2024 12:15:38	U	SMPL