

Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Iwona

Parameter: COD

SUPERVISOR REVIEW BY: Sohil

Run Number: LB129263

Reagent/Standard	Lot/Log #
COD calibration std. 10 ppm	WP105791
COD calibration std. 150 ppm	WP105794
COD calibration std. 50 ppm	WP105792
COD calibration std. 0 ppm	WP105790
COD Digestion Vials Low Level 0-150Mg/L	W3066
COD ICV-LCS std, 50ppm	WP105920
COD calibration std. 100 ppm	WP105793
COD ICV-LCS std, 50ppm	WP106323
COD CCV std, 50ppm	WP106322
COD LOD std, 5ppm	WP106324
COD LOQ std, 10.0PPM	WP106325

Temp In(C): 148	Date In: 01/26/2024	Time In: 09:15
Temp Out(C): 151	Date Out: 01/26/2024	Time Out: 11:15

Intercept: 0.0126

Slope: 1.003

Regression: 0.9999

Seq	Lab ID	TrueValue (mg/l)	DF	MATRIX	Reading	Result (mg/l)	%D	Anal Date	Anal Time
1	CAL1	0	1	Water	0.000	-0.013		12/19/2023	14:30
2	CAL2	10	1	Water	11.000	10.955	9.6	12/19/2023	14:30
3	CAL3	50	1	Water	49.000	48.841	-2.3	12/19/2023	14:31
4	CAL4	100	1	Water	100.000	99.688	-0.3	12/19/2023	14:31
5	CAL5	150	1	Water	151.000	150.536	0.4	12/19/2023	14:32



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Iwona

Parameter: COD

SUPERVISOR REVIEW BY: Sohil

Run Number: LB129263

Seq	Lab ID	True Value (mg/l)	Initial Weight (g)	Final Vol (ml)	DF	MATRIX	Reading	Result	AnalDate	AnalTime
1	ICV	50	NA	NA	1	Water	50.000	49.838	12/19/2023	14:32
2	ICB		NA	NA	1	Water	1.000	0.984	12/19/2023	14:33
3	CCV1	50	NA	NA	1	Water	51.000	50.835	01/26/2024	13:20
4	CCB1		NA	NA	1	Water	1.000	0.984	01/26/2024	13:20
5	LB129263BL		NA	NA	1	Water	1.000	0.984	01/26/2024	13:21
6	LB129263BS	50	NA	NA	1	Water	51.000	50.835	01/26/2024	13:21
7	P1163-01		NA	NA	10	Water	37.000	36.877	01/26/2024	13:22
8	P1163-01DUP		NA	NA	10	Water	36.000	35.880	01/26/2024	13:22
9	P1163-01MS	50	NA	NA	10	Water	41.000	40.865	01/26/2024	13:23
10	P1163-01MSD	50	NA	NA	10	Water	41.000	40.865	01/26/2024	13:23
11	P1166-01		NA	NA	10	Water	39.000	38.871	01/26/2024	13:24
12	P1187-07		NA	NA	1	Water	5.000	4.972	01/26/2024	13:24
13	P1187-08		NA	NA	1	Water	10.000	9.958	01/26/2024	13:25
14	CCV2	50	NA	NA	1	Water	50.000	49.838	01/26/2024	13:25
15	CCB2		NA	NA	1	Water	1.000	0.984	01/26/2024	13:26