



Analytical Summary Report

Analysis Method: SM5220 D
Parameter: COD
Run Number: LB129266

ANALYST: Iwona
SUPERVISOR REVIEW BY: Sohil

Table with 2 columns: Reagent/Standard, Lot/Log #. Rows include COD calibration standards (10, 150, 50, 0 ppm), COD Digestion Vials, and COD ICV-LCS standards (50, 100 ppm).

Table with 3 columns: Temp In (C), Date In, Time In; Temp Out (C), Date Out, Time Out. Values: Temp In: 148, Date In: 01/26/2024, Time In: 09:15; Temp Out: 151, Date Out: 01/26/2024, Time Out: 11:15.

Intercept: 0.0126 Slope: 1.003 Regression: 0.9999

Table with 10 columns: Seq, Lab ID, TrueValue (mg/l), DF, MATRIX, Reading, Result (mg/l), %D, Anal Date, Anal Time. Contains 5 rows of calibration data (CAL1 to CAL5).



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Iwona

Parameter: COD

SUPERVISOR REVIEW BY: Sohil

Run Number: LB129266

Seq	Lab ID	True Value (mg/l)	Initial Weight (g)	Final Vol (ml)	DF	MATRIX	Reading	Result	AnalDate	AnalTime
1	ICV	50	NA	NA	1	Water	50.000	49.838	12/19/2023	14:32
2	ICB		NA	NA	1	Water	1.000	0.984	12/19/2023	14:33
3	CCV1	50	NA	NA	1	Water	49.000	48.841	01/26/2024	13:30
4	CCB1		NA	NA	1	Water	1.000	0.984	01/26/2024	13:30
5	LB129266BL		1	50	1	Solid	0.000	-0.013	01/26/2024	13:31
6	LB129266BS	50	1	50	1	Solid	50.000	49.838	01/26/2024	13:31
7	P1187-01		1	50	1	Solid	5.000	4.972	01/26/2024	13:32
8	P1187-02		1	50	1	Solid	11.000	10.955	01/26/2024	13:32
9	CCV2	50	NA	NA	1	Water	50.000	49.838	01/26/2024	13:33
10	CCB2		NA	NA	1	Water	1.000	0.984	01/26/2024	13:33