



Analytical Summary Report

Analysis Method: 9040C
Parameter: pH
Run Number: LB129268

Analyst By : jignesh
Supervisor Review By : Iwona
Slope : 99.2
pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	01/26/2024	13:45
2	CAL2	1	Water	NA	NA	20.3	7.00	01/26/2024	13:46
3	CAL3	1	Water	NA	NA	20.4	10.02	01/26/2024	13:50
4	ICV	1	Water	NA	NA	20.3	7.01	01/26/2024	13:55
5	CCV1	1	Water	NA	NA	20.2	2.01	01/26/2024	15:00
6	P1264-03	1	Water	NA	NA	20.2	6.12	01/26/2024	15:05
7	P1264-03DUP	1	Water	NA	NA	20.3	6.13	01/26/2024	15:06
8	CCV2	1	Water	NA	NA	20.3	12.02	01/26/2024	15:10

JP 121268

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH P1264 WorkList ID : 177428 Department : Wet-Chemistry Date : 01-26-2024 13:37:57

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1264-03	17 257221	Water	pH	Cool 4 deg C	PSEG03	E51	01/26/2024	9040C

Date/Time 01.26.24 13:40
Raw Sample Received by: SA (col C)
Raw Sample Relinquished by: JTC (SM)

Date/Time 01.26.24 16:30
Raw Sample Received by: JTC (SM)
Raw Sample Relinquished by: SA (col C)