

LB129269


SFAM01.1

INSTRUMENT ID: CV1

Linear

μ Abs.:

51853



A= 0.0000e+000

B= 1.9333e-004 slope

C= 5.2644e-002 y-intercept

Rho= 0.9998611

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.071	0.071	94	0.000	94					
0.05	0.050					442	0				
0.20	0.200	0.209	0.009	811	0.0 %	811					
2.50	2.500	2.473	-0.027	12520	0.0 %	12520					
5.00	5.000	4.900	-0.100	25074	0.0 %	25074					
7.50	7.500	7.469	-0.031	38364	0.0 %	38364					
10.0	10.000	10.077	0.077	51853	0.0 %	51853					

0.00
0.05
0.20
2.50
5.00
7.50
10.0

LB129269

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Stnd Conc	Method	Units	Date	Type
	0 S0	94	-		0 SFAM01.1	PPB	1/26/2024 15:42	Std
	0.2 S01	811	-		0.2 SFAM01.1	PPB	1/26/2024 15:45	Std
	2.5 S02	12520	-		2.5 SFAM01.1	PPB	1/26/2024 15:47	Std
	5 S03	25074	-		5 SFAM01.1	PPB	1/26/2024 15:49	Std
	7.5 S04	38364	-		7.5 SFAM01.1	PPB	1/26/2024 15:56	Std
	10 S05	51853	-		10 SFAM01.1	PPB	1/26/2024 15:58	Std
ICV082	ICV082	19341	3.7918 -		SFAM01.1	PPB	1/26/2024 16:01	SMPL
ICB082	ICB082	-156	0.0225 -		SFAM01.1	PPB	1/26/2024 16:03	SMPL
CCV028	CCV028	25391	4.9614 -		SFAM01.1	PPB	1/26/2024 16:05	SMPL
CCB028	CCB028	-154	0.0229 -		SFAM01.1	PPB	1/26/2024 16:08	SMPL
PB158667BL	PBW667	220	0.0952 -		SFAM01.1	PPB	1/26/2024 16:10	SMPL
PB158610TB	LEB610	-28	0.0472 -		SFAM01.1	PPB	1/26/2024 16:12	SMPL
P1195-01	E0X65	176	0.0867 -		SFAM01.1	PPB	1/26/2024 16:14	SMPL
P1195-02	E0X66	188	0.089 -		SFAM01.1	PPB	1/26/2024 16:17	SMPL
P1195-03	E0X66D	158	0.0832 -		SFAM01.1	PPB	1/26/2024 16:19	SMPL
P1195-04	E0X66S	4396	0.9025 -		SFAM01.1	PPB	1/26/2024 16:21	SMPL
P1195-05	E0XF0	1334	0.3105 -		SFAM01.1	PPB	1/26/2024 16:23	SMPL
P1195-06	E0Y20	338	0.118 -		SFAM01.1	PPB	1/26/2024 16:26	SMPL
CCV029	CCV029	24743	4.8361 -		SFAM01.1	PPB	1/26/2024 16:28	SMPL
CCB029	CCB029	-154	0.0229 -		SFAM01.1	PPB	1/26/2024 16:30	SMPL