

LB129279


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

43213



A= 0.0000e+000

B= 2.3055e-004

C= 3.5106e-002

Rho= 0.9999054

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.059	0.059	104	0.000	104	0			
0.2	0.200	0.197	-0.003	701	0.0 %	701				
2.5	2.500	2.456	-0.044	10500	0.0 %	10500				
5.0	5.000	4.925	-0.075	21208	0.0 %	21208				
7.5	7.500	7.566	0.066	32664	0.0 %	32664				
10.0	10.000	9.998	-0.002	43213	0.0 %	43213				

6/21
12-21-0

LB129279 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 29 Jan 2024 09:58:25

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	104	-	0.0000	7470A	PPB	29 Jan 2024 10:26:23	S	Std
0.2 - 1	10.2	701	-	0.2000	7470A	PPB	29 Jan 2024 10:28:40	S	Std
2.5 - 1	125	10500	-	2.5000	7470A	PPB	29 Jan 2024 10:30:57	S	Std
5.0 - 1	55	21208	-	5.0000	7470A	PPB	29 Jan 2024 10:33:14	S	Std
7.5 - 1	57.5	32664	-	7.5000	7470A	PPB	29 Jan 2024 10:38:32	S	Std
10.0 - 1	510	43213	-	10.0000	7470A	PPB	29 Jan 2024 10:40:48	S	Std
ICV83 - 1	ICV83	17870	4.1550	-7470A	PPB	29 Jan 2024 10:43:42	U	SMPL	
ICB83 - 1	ICB83	-277	-0.0288	-7470A	PPB	29 Jan 2024 10:45:57	U	SMPL	
CCV30 - 1	CCV30	21609	5.0171	-7470A	PPB	29 Jan 2024 10:48:16	U	SMPL	
CCB30 - 1	CCB30	-280	-0.0294	-7470A	PPB	29 Jan 2024 10:50:32	U	SMPL	
CRA - 1	CRA	586	0.1702	-7470A	PPB	29 Jan 2024 10:52:49	U	SMPL	
HighStd - 1	HighStd	43523	10.0694	-7470A	PPB	29 Jan 2024 10:55:05	U	SMPL	
ChkStd - 1	ChkStd	28860	6.6888	-7470A	PPB	29 Jan 2024 10:57:21	U	SMPL	
PB158631BL - 1	PBW	-396	-0.0562	-7470A	PPB	29 Jan 2024 10:59:40	U	SMPL	
PB158631BS - 1	LCSW	16361	3.8071	-7470A	PPB	29 Jan 2024 11:01:59	U	SMPL	
P1240-03 - 1	72-11985	-248	-0.0221	-7470A	PPB	29 Jan 2024 11:04:16	U	SMPL	
P1240-06 - 1	VNJ-232	-37	0.0266	-7470A	PPB	29 Jan 2024 11:06:34	U	SMPL	
P1240-09 - 1	VNJ-223	-27	0.0289	-7470A	PPB	29 Jan 2024 11:08:51	U	SMPL	
P1243-02 - 1	MR-308-0-2	31	0.0423	-7470A	PPB	29 Jan 2024 11:11:09	U	SMPL	
P1243-04 - 1	MR-308-17-19	30	0.0420	-7470A	PPB	29 Jan 2024 11:13:27	U	SMPL	
CCV31 - 1	CCV31	21476	4.9864	-7470A	PPB	29 Jan 2024 11:15:47	U	SMPL	
CCB31 - 1	CCB31	-234	-0.0188	-7470A	PPB	29 Jan 2024 11:18:08	U	SMPL	
P1246-03 - 1	TP-COMP-012424	33	0.0427	-7470A	PPB	29 Jan 2024 11:20:26	U	SMPL	
P1246-03DUP - 1	TP-COMP-012424DUP	17	0.0390	-7470A	PPB	29 Jan 2024 11:22:41	U	SMPL	
P1246-03MS - 1	TP-COMP-012424MS	15897	3.7002	-7470A	PPB	29 Jan 2024 11:24:57	U	SMPL	
P1246-03MSD - 1	TP-COMP-012424MSD	17142	3.9872	-7470A	PPB	29 Jan 2024 11:27:13	U	SMPL	
PB158616TB - 1	PB158616TB	-231	-0.0182	-7470A	PPB	29 Jan 2024 11:29:32	U	SMPL	
P1246-03LX5 - 1		7	0.0367	-7470A	PPB	29 Jan 2024 11:31:50	U	SMPL	
P1246-03A - 1		17489	4.0672	-7470A	PPB	29 Jan 2024 11:34:12	U	SMPL	
CCV32 - 1	CCV32	21929	5.0908	-7470A	PPB	29 Jan 2024 11:36:30	U	SMPL	
CCB32 - 1	CCB32	-243	-0.0209	-7470A	PPB	29 Jan 2024 11:38:49	U	SMPL	