CHEMTECH

Analytical Summary Report

Analysis Method: 9040C
Parameter: Corrosivity
Run Number: LB129280

Analyst By : jignesh
Supervisor Review By : Iwona
Slope : 99.2
pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	01/29/2024	09:25
2	CAL2	1	Water	NA	NA	20.3	7.00	01/29/2024	09:26
3	CAL3	1	Water	NA	NA	20.3	10.02	01/29/2024	09:28
4	ICV	1	Water	NA	NA	20.4	7.02	01/29/2024	09:35
5	CCV1	1	Water	NA	NA	20.3	2.01	01/29/2024	09:37
6	P1270-04	1	Water	NA	NA	20.1	9.13	01/29/2024	09:45
7	P1270-04DUP	1	Water	NA	NA	20.2	9.14	01/29/2024	09:46
8	CCV2	1	Water	NA	NA	20.3	12.02	01/29/2024	09:50

VB 129280

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity p1270 water **WorkList ID :** 177444 **Department :** Wet-Chemistry **Date :** 01-29-2024 09:14:33

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1270-04 C	WC-LIQUID-20240126	Water	Corrosivity	Cool 4 deg C	PARS02	D21	01/26/2024	9040C

Date/Time 01.29.24 09:20
Raw Sample Received by: JB (CDC)
Raw Sample Relinquished by: CR SM

Date/Time 01.29.24 11:30
Raw Sample Received by: CR SM
Raw Sample Relinquished by: JB (CDC)