



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB129281

Slope : 99.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	01/29/2024	09:25
2	CAL2	1	Water	NA	NA	20.3	7.00	01/29/2024	09:26
3	CAL3	1	Water	NA	NA	20.3	10.02	01/29/2024	09:28
4	ICV	1	Water	NA	NA	20.4	7.02	01/29/2024	09:35
5	CCV1	1	Water	NA	NA	20.4	2.01	01/29/2024	09:55
6	P1270-02	1	Solid	20.02	20	23.9	8.60	01/29/2024	10:10
7	P1270-02DUP	1	Solid	20.03	20	23.8	8.61	01/29/2024	10:11
8	CCV2	1	Water	NA	NA	20.4	12.02	01/29/2024	10:15

✓ 129281

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity p1270 WorkList ID : 177445 Department : Wet-Chemistry Date : 01-29-2024 09:15:28

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1270-02	WC-SOIL-20240126	Solid	Corrosivity	Cool 4 deg C	PARS02	D21	01/26/2024	9045D

Date/Time 01-29-24 09:20
Raw Sample Received by: 701001
Raw Sample Relinquished by: CB SM

Date/Time 01-29-24 11:30
Raw Sample Received by: CP SM
Raw Sample Relinquished by: 701001