



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB129282

Slope : 99.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	01/29/2024	09:30
2	CAL2	1	Water	NA	NA	20.2	7.00	01/29/2024	09:31
3	CAL3	1	Water	NA	NA	20.2	10.02	01/29/2024	09:33
4	ICV	1	Water	NA	NA	20.5	7.01	01/29/2024	09:35
5	CCV1	1	Water	NA	NA	20.3	2.01	01/29/2024	09:44
6	P1217-01	1	Solid	20.02	20	23.4	10.28	01/29/2024	09:50
7	P1217-02	1	Solid	20.02	20	23.4	10.28	01/29/2024	09:50
8	P1217-03	1	Solid	20.02	20	23.4	10.28	01/29/2024	09:50
9	CCV2	1	Water	NA	NA	20.3	12.02	01/29/2024	10:00

WORKLIST(Hardcopy Internal Chain)

13129282

WorkList Name : ph p1217 WorkList ID : 177443 Department : Wet-Chemistry Date : 01-29-2024 09:13:31

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1217-01	E0X61	Solid	pH	Cool 4 deg C	USEP01	Q52	01/18/2024	9045D
P1217-02	E0X61D	Solid	pH	Cool 4 deg C	USEP01	Q52	01/18/2024	9045D
P1217-03	E0X61S	Solid	pH	Cool 4 deg C	USEP01	Q52	01/18/2024	9045D

Date/Time 01-29-24 09:20

Raw Sample Received by: JALC/CCL

Raw Sample Relinquished by: CP SM

Date/Time 01-29-24 11:30

Raw Sample Received by: CP SM

Raw Sample Relinquished by: JALC/CCL