



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB129294

Slope : 98.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	01/29/2024	12:45
2	CAL2	1	Water	NA	NA	20.2	6.99	01/29/2024	12:46
3	CAL3	1	Water	NA	NA	20.3	10.02	01/29/2024	12:48
4	ICV	1	Water	NA	NA	20.3	7.01	01/29/2024	12:50
5	CCV1	1	Water	NA	NA	20.3	2.01	01/29/2024	13:00
6	P1272-01	1	Water	NA	NA	20.2	4.96	01/29/2024	13:10
7	P1272-01DUP	1	Water	NA	NA	20.2	4.97	01/29/2024	13:11
8	P1272-02	1	Water	NA	NA	20.2	4.82	01/29/2024	13:15
9	P1272-03	1	Water	NA	NA	20.1	5.14	01/29/2024	13:20
10	CCV2	1	Water	NA	NA	20.3	12.02	01/29/2024	13:25

12/29/24

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH P1272

WorkList ID : 177462

Department : Wet-Chemistry

Date : 01-29-2024 12:32:55

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1272-01	A	Water	pH	Cool 4 deg C	STER09	D11	01/26/2024	9040C
P1272-02	A GEMINI	Water	pH	Cool 4 deg C	STER09	D11	01/26/2024	9040C
P1272-03	A D	Water	pH	Cool 4 deg C	STER09	D11	01/26/2024	9040C

Date/Time 01/29/24 12:40
 Raw Sample Received by: RD/WDJ
 Raw Sample Relinquished by: 12/29/24

Date/Time 01/29/24 15:00
 Raw Sample Received by: 12/29/24
 Raw Sample Relinquished by: 12/29/24