



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB129303

Slope : 99.3

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	01/29/2024	17:00
2	CAL2	1	Water	NA	NA	20.3	6.99	01/29/2024	17:01
3	CAL3	1	Water	NA	NA	20.2	10.02	01/29/2024	17:05
4	ICV	1	Water	NA	NA	20.4	7.02	01/29/2024	17:10
5	CCV1	1	Water	NA	NA	20.4	2.01	01/29/2024	17:11
6	P1276-02	1	Solid	20.02	20	22.3	6.90	01/29/2024	17:15
7	P1276-04	1	Solid	20.03	20	21.9	7.35	01/29/2024	17:19
8	P1276-06	1	Solid	20.05	20	22.4	7.33	01/29/2024	17:22
9	P1277-02	1	Solid	20.02	20	21.7	7.64	01/29/2024	17:25
10	P1277-04	1	Solid	20.03	20	22.5	7.23	01/29/2024	17:26
11	P1277-06	1	Solid	20.04	20	22.2	7.54	01/29/2024	17:30
12	P1277-06DUP	1	Solid	20.03	20	22.3	7.55	01/29/2024	17:31
13	CCV2	1	Water	NA	NA	20.3	12.02	01/29/2024	17:33

829303

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosion p1280

WorkList ID : 177469

Department : Wet-Chemistry

Date : 01-29-2024 16:43:28

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1276-02	TP-36	Solid	Corrosivity	Cool 4 deg C	PSEG03	D11	01/29/2024	9045D
P1276-04	TP-37	Solid	Corrosivity	Cool 4 deg C	PSEG03	D11	01/29/2024	9045D
P1276-06	TP-38	Solid	Corrosivity	Cool 4 deg C	PSEG03	D11	01/29/2024	9045D
P1277-02	TP-39	Solid	Corrosivity	Cool 4 deg C	PSEG03	D11	01/29/2024	9045D
P1277-04	TP-40	Solid	Corrosivity	Cool 4 deg C	PSEG03	D11	01/29/2024	9045D
P1277-06	TP-41	Solid	Corrosivity	Cool 4 deg C	PSEG03	D11	01/29/2024	9045D

Date/Time 012924 16:50

Raw Sample Received by: JG (WC)

Raw Sample Relinquished by: JTC (W)

Date/Time 01-29-24

Raw Sample Received by: JTC (W)

Raw Sample Relinquished by: JG (WC)

Page 1 of 1