



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB129306

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	01/30/2024	09:25
2	CAL2	1	Water	NA	NA	20.3	7.00	01/30/2024	09:26
3	CAL3	1	Water	NA	NA	20.5	10.02	01/30/2024	09:30
4	ICV	1	Water	NA	NA	20.5	7.02	01/30/2024	09:35
5	CCV1	1	Water	NA	NA	20.3	2.01	01/30/2024	09:40
6	P1278-01	1	Water	NA	NA	22.8	7.70	01/30/2024	09:45
7	P1278-01DUP	1	Water	NA	NA	22.9	7.71	01/30/2024	09:46
8	P1278-03	1	Water	NA	NA	23.3	8.11	01/30/2024	09:50
9	CCV2	1	Water	NA	NA	20.3	12.02	01/30/2024	10:00