

LB129307


7470A

INSTRUMENT ID: CYL

Linear

 μ Abs.:

43761



A= 0.0000e+000

B= 2.2845e-004

C= 2.7565e-002

Rho= 0.9998957

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	n/SD
0.0	0.000	0.060	0.060	144	0.000	144					
0.2	0.200	0.209	0.009	793	0.0 %	793	0				4
2.5	2.500	2.461	-0.039	10651	0.0 %	10651					2
5.0	5.000	4.903	-0.097	21342	0.0 %	21342					2
7.5	7.500	7.542	0.042	32894	0.0 %	32894					1
10.0	10.000	10.025	0.025	43761	0.0 %	43761					0

LB129307 INSTRUMENT ID :: CV1

Method: 7470A

Operator: Admin

Date of Analysis: 30 Jan 2024 11:24:10

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	144	-	0.0000	7470A	PPB	30 Jan 2024 11:43:59	S	Std
0.2 - 1	502	793	-	0.2000	7470A	PPB	30 Jan 2024 11:46:16	S	Std
2.5 - 1	52.5	10651	-	2.5000	7470A	PPB	30 Jan 2024 11:48:33	S	Std
5.0 - 1	55	21342	-	5.0000	7470A	PPB	30 Jan 2024 11:50:50	S	Std
7.5 - 1	57.5	32894	-	7.5000	7470A	PPB	30 Jan 2024 11:53:08	S	Std
10.0 - 1	511	43761	-	10.0000	7470A	PPB	30 Jan 2024 11:58:04	S	Std
ICV85 - 1	ICV85	16895	3.8872	-	7470A	PPB	30 Jan 2024 12:03:44	U	SMPL
ICB85 - 1	ICB85	79	0.0456	-	7470A	PPB	30 Jan 2024 12:05:59	U	SMPL
CCV36 - 1	CCV36	20916	4.8058	-	7470A	PPB	30 Jan 2024 12:08:16	U	SMPL
CCB36 - 1	CCB36	50	0.0390	-	7470A	PPB	30 Jan 2024 12:10:31	U	SMPL
CRA - 1	CRA	696	0.1866	-	7470A	PPB	30 Jan 2024 12:12:49	U	SMPL
HighStd - 1	HighStd	44439	10.1796	-	7470A	PPB	30 Jan 2024 12:15:04	U	SMPL
ChkStd - 1	ChkStd	30792	7.0620	-	7470A	PPB	30 Jan 2024 12:17:18	U	SMPL
PB158685BL - 1	PBW	-155	-0.0078	-	7470A	PPB	30 Jan 2024 12:19:41	U	SMPL
PB158685BS - 1	LCSW	18096	4.1616	-	7470A	PPB	30 Jan 2024 12:22:01	U	SMPL
P1258-04 - 1	S-1	2073	0.5011	-	7470A	PPB	30 Jan 2024 12:24:17	U	SMPL
P1270-02 - 1	WC-SOIL-20240126	591	0.1626	-	7470A	PPB	30 Jan 2024 12:26:35	U	SMPL
P1270-02DUP - 1	WC-SOIL-20240126DUP	405	0.1201	-	7470A	PPB	30 Jan 2024 12:28:54	U	SMPL
P1270-02MS - 1	WC-SOIL-20240126MS	18932	4.3526	-	7470A	PPB	30 Jan 2024 12:31:11	U	SMPL
P1270-02MSD - 1	WC-SOIL-20240126MSD	19428	4.4659	-	7470A	PPB	30 Jan 2024 12:36:09	U	SMPL
CCV37 - 1	CCV37	21182	4.8666	-	7470A	PPB	30 Jan 2024 12:38:24	U	SMPL
CCB37 - 1	CCB37	177	0.0680	-	7470A	PPB	30 Jan 2024 12:40:42	U	SMPL
PB158686BL - 1	PBW	256	0.0860	-	7470A	PPB	30 Jan 2024 12:43:00	U	SMPL
PB158686BS - 1	LCSW	15638	3.6001	-	7470A	PPB	30 Jan 2024 12:45:15	U	SMPL
P1270-04 - 1	WC-LIQUID-20240126	84	0.0468	-	7470A	PPB	30 Jan 2024 12:47:31	U	SMPL
P1270-04DUP - 1	WC-LIQUID-20240126DUP	248	0.0842	-	7470A	PPB	30 Jan 2024 12:49:49	U	SMPL
P1270-04MS - 1	WC-LIQUID-20240126MS	16419	3.7785	-	7470A	PPB	30 Jan 2024 12:52:05	U	SMPL
P1270-04MSD - 1	WC-LIQUID-20240126MSD	17920	4.1214	-	7470A	PPB	30 Jan 2024 12:54:22	U	SMPL
PB158707BL - 1	PBW	-41	0.0182	-	7470A	PPB	30 Jan 2024 12:56:42	U	SMPL
PB158707BS - 1	LCSW	18507	4.2555	-	7470A	PPB	30 Jan 2024 12:59:01	U	SMPL
P1264-03 - 1	257221	2796	0.6663	-	7470A	PPB	30 Jan 2024 13:01:18	U	SMPL
P1264-03DUP - 1	257221DUP	2639	0.6304	-	7470A	PPB	30 Jan 2024 13:03:36	U	SMPL
CCV38 - 1	CCV38	21186	4.8675	-	7470A	PPB	30 Jan 2024 13:05:53	U	SMPL
CCB38 - 1	CCB38	82	0.0463	-	7470A	PPB	30 Jan 2024 13:08:10	U	SMPL
P1264-03MS - 1	257221MS	18818	4.3265	-	7470A	PPB	30 Jan 2024 13:10:29	U	SMPL
P1264-03MSD - 1	257221MSD	20761	4.7704	-	7470A	PPB	30 Jan 2024 13:12:44	U	SMPL
P1270-03 - 1	WC-LIQUID-20240126	130	0.0573	-	7470A	PPB	30 Jan 2024 13:15:03	U	SMPL
PB158653TB - 1	PB158653TB	249	0.0844	-	7470A	PPB	30 Jan 2024 13:17:21	U	SMPL
PB158668TB - 1	PB158668TB	267	0.0886	-	7470A	PPB	30 Jan 2024 13:19:37	U	SMPL
CCV39 - 1	CCV39	21306	4.8949	-	7470A	PPB	30 Jan 2024 13:21:53	U	SMPL
CCB39 - 1	CCB39	58	0.0408	-	7470A	PPB	30 Jan 2024 13:24:10	U	SMPL
P1270-02LX5 - 1		128	0.0568	-	7470A	PPB	30 Jan 2024 13:26:30	U	SMPL
P1270-04LX5 - 1		-2	0.0271	-	7470A	PPB	30 Jan 2024 13:31:03	U	SMPL
P1270-04A - 1		18165	4.1773	-	7470A	PPB	30 Jan 2024 13:33:18	U	SMPL
P1264-03LX5 - 1		437	0.1274	-	7470A	PPB	30 Jan 2024 13:35:34	U	SMPL
P1264-03A - 1		20416	4.6916	-	7470A	PPB	30 Jan 2024 13:37:52	U	SMPL
P1270-02A - 1		16352	3.7632	-	7470A	PPB	30 Jan 2024 13:48:20	U	SMPL
CCV40 - 1	CCV40	22106	5.0777	-	7470A	PPB	30 Jan 2024 13:50:51	U	SMPL
CCB40 - 1	CCB40	-172	-0.0117	-	7470A	PPB	30 Jan 2024 13:53:08	U	SMPL