



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Nikita

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB129309

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2745

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.4	01/29/2024	08:30
2	P1235-10		1	14	100.00	>212.0	>212.0	01/29/2024	09:00
3	P1235-10DUP		1	13	100.00	>212.0	>212.0	01/29/2024	09:30
4	P1256-01		1	12	100.00	>212.0	>212.0	01/29/2024	10:00
5	P1258-01		1	13	100.00	>212.0	>212.0	01/29/2024	10:30
6	P1258-02		1	12	100.00	>212.0	>212.0	01/29/2024	11:00
7	P1258-03		1	11	100.00	>212.0	>212.0	01/29/2024	11:30
8	P1264-03		1	14	100.00	>212.0	>212.0	01/29/2024	12:00
9	P1270-04		1	14	100.00	>212.0	>212.0	01/29/2024	12:30

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB 129309

WORKLIST(Hardcopy Internal Chain)

WorkList Name : flashpoint 01/29

WorkList ID : 177459

Department : Wet-Chemistry

Date : 01-29-2024 08:10:38

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1235-10	1107	Water	Flash Point	Cool 4 deg C	PSEG03	E41	01/23/2024	1010B
P1256-01	60131	Water	Flash Point	Cool 4 deg C	PSEG03	E51	01/25/2024	1010B
P1258-01	50234	Water	Flash Point	Cool 4 deg C	PSEG03	E41	01/25/2024	1010B
P1258-02	50235	Water	Flash Point	Cool 4 deg C	PSEG03	E41	01/25/2024	1010B
P1258-03	50236	Water	Flash Point	Cool 4 deg C	PSEG03	E41	01/25/2024	1010B
P1264-03	257221	Water	Flash Point	Cool 4 deg C	PSEG03	E51	01/26/2024	1010B
P1270-04	WC-LIQUID-20240126	Water	Flash Point	Cool 4 deg C	PARS02	D21	01/26/2024	1010B

Date/Time 01/29/24 8:20
Raw Sample Received by: N.P. W.C.
Raw Sample Relinquished by: [Signature]

Date/Time 01/29/24 13:45
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: N.P. W.C.