



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB129313

Slope : 97.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	01/31/2024	08:10
2	CAL2	1	Water	NA	NA	20.3	7.00	01/31/2024	08:11
3	CAL3	1	Water	NA	NA	20.3	10.02	01/31/2024	08:15
4	ICV	1	Water	NA	NA	20.4	7.02	01/31/2024	08:20
5	CCV1	1	Water	NA	NA	20.3	2.02	01/31/2024	08:25
6	P1291-01	1	Water	NA	NA	20.3	4.83	01/31/2024	08:30
7	P1291-01DUP	1	Water	NA	NA	20.4	4.84	01/31/2024	08:31
8	CCV2	1	Water	NA	NA	20.4	12.02	01/31/2024	08:35

WB 129313

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph p1291 WorkList ID : 177499 Department : Wet-Chemistry Date : 01-31-2024 07:50:35

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1291-01	✓ HAW-RIVER-MAIN-CELL	Water	pH	Cool 4 deg C	STER11	D11	01/24/2024	9040C

01-31-24

Date/Time 01-31-23 08:00
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]

Date/Time 01-31-24 10:00
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]