


LB129314

7471B

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	
0.0	0.000	-0.006	-0.006	194	0.000	194					
0.2	0.200	0.219	0.019	1556	0.0 %	1556					
2.5	2.500	2.424	-0.076	14905	0.0 %	14905					
5.0	5.000	5.012	0.012	30569	0.0 %	30569					
7.5	7.500	7.674	0.174	46686	0.0 %	46686					
10.0	10.000	9.877	-0.123	60020	0.0 %	60020					

4/10
19
-3
0
2
-1

Method: 7471B

Operator: Admin

LB129314 INSTRUMENT ID : CV1

Date of Analysis: 31 Jan 2024 10:55:15

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	T
0.0 - 1	50	194	-	0.000	7471B	PPB	31 Jan 2024 11:13:04	S	S
0.2 - 1	50.2	1556	-	0.200	7471B	PPB	31 Jan 2024 11:15:22	S	S
2.5 - 1	52.5	14905	-	2.500	7471B	PPB	31 Jan 2024 11:17:39	S	S
5.0 - 1	55	30569	-	5.000	7471B	PPB	31 Jan 2024 11:19:55	S	S
7.5 - 1	57.5	46686	-	7.500	7471B	PPB	31 Jan 2024 11:22:15	S	S
10.0 - 1	50	60020	-	10.000	7471B	PPB	31 Jan 2024 11:27:26	S	S
ICV86 - 1	ICV86	25988	4.2550	-7471B	PPB	31 Jan 2024 11:30:35	U	SM	
ICB86 - 1	ICB86	-115	-0.0570	-7471B	PPB	31 Jan 2024 11:32:52	U	SM	
CCV41 - 1	CCV41	31585	5.1796	-7471B	PPB	31 Jan 2024 11:35:17	U	SM	
CCB41 - 1	CCB41	-145	-0.0620	-7471B	PPB	31 Jan 2024 11:37:31	U	SM	
CRA - 1	CRA	1521	0.2132	-7471B	PPB	31 Jan 2024 11:39:55	U	SM	
HighStd - 1	HighStd	60049	9.8817	-7471B	PPB	31 Jan 2024 11:42:10	U	SM	
ChkStd - 1	ChkStd	45336	7.4512	-7471B	PPB	31 Jan 2024 11:44:27	U	SM	
PB158731BL - 1	PBS	-311	-0.0894	-7471B	PPB	31 Jan 2024 11:46:55	U	SM	
PB158731BS - 1	LCSS	25401	4.1580	-7471B	PPB	31 Jan 2024 11:49:23	U	SM	
P1276-01 - 1	TP-36	92974	15.3207	-7471B	PPB	31 Jan 2024 11:51:40	U	SM	
P1276-03 - 1	TP-37	393189	HIGH	-7471B	PPB	31 Jan 2024 11:54:02	X	SM	
P1276-05 - 1	TP-38	42235	6.9389	-7471B	PPB	31 Jan 2024 11:56:14	U	SM	
P1277-01 - 1	TP-39	19978	3.2622	-7471B	PPB	31 Jan 2024 11:58:58	U	SM	
P1277-03 - 1	TP-40	237307	39.1636	-7471B	PPB	31 Jan 2024 12:01:24	U	SM	
CCV42 - 1	CCV42	30908	5.0678	-7471B	PPB	31 Jan 2024 12:03:44	U	SM	
CCB42 - 1	CCB42	-350	-0.0959	-7471B	PPB	31 Jan 2024 12:06:22	U	SM	
P1277-05 - 1	TP-41	14500	2.3573	-7471B	PPB	31 Jan 2024 12:08:51	U	SM	
P1280-01 - 1	Manhole	3077	0.4703	-7471B	PPB	31 Jan 2024 12:11:06	U	SM	
P1283-01 - 1	OR-02-012924	4457	0.6982	-7471B	PPB	31 Jan 2024 12:13:25	U	SM	
P1285-01 - 1	SECOND-AVE-TP-1	2399	0.3583	-7471B	PPB	31 Jan 2024 12:15:42	U	SM	
P1285-03 - 1	CHESTNUT-ST-TP-2	2175	0.3213	-7471B	PPB	31 Jan 2024 12:22:34	U	SM	
P1285-03DUP - 1	CHESTNUT-ST-TP-2DUP	2198	0.3251	-7471B	PPB	31 Jan 2024 12:24:50	U	SM	
P1285-03MS - 1	CHESTNUT-ST-TP-2MS	25211	4.1267	-7471B	PPB	31 Jan 2024 12:27:07	U	SM	
P1285-03MSD - 1	CHESTNUT-ST-TP-2MSD	26527	4.3440	-7471B	PPB	31 Jan 2024 12:35:20	U	SM	
P1295-01 - 1	OK-02-013024	1065	0.1379	-7471B	PPB	31 Jan 2024 12:37:37	U	SM	
P1296-01 - 1	EO-01-013024	2421	0.3619	-7471B	PPB	31 Jan 2024 12:40:04	U	SM	
CCV43 - 1	CCV43	32229	5.2860	-7471B	PPB	31 Jan 2024 12:42:21	U	SM	
CCB43 - 1	CCB43	-174	-0.0668	-7471B	PPB	31 Jan 2024 12:44:38	U	SM	
P1297-01 - 1	TR-05-013024	3117	0.4769	-7471B	PPB	31 Jan 2024 12:47:01	U	SM	
P1270-01DLX5 - 1		33730	5.5339	-7471B	PPB	31 Jan 2024 12:49:16	U	SM	
P1285-03A - 1		27535	4.5106	-7471B	PPB	31 Jan 2024 12:53:53	U	SM	
P1276-01DLX5 - 1		18605	3.0354	-7471B	PPB	31 Jan 2024 12:56:09	U	SM	
P1276-03DLX10 - 1		56923	9.3653	-7471B	PPB	31 Jan 2024 12:58:32	U	SM	
P1277-03DLX10 - 1		27693	4.5367	-7471B	PPB	31 Jan 2024 13:04:00	U	SM	
P1285-03LX5 - 1		488	0.0426	-7471B	PPB	31 Jan 2024 13:06:16	U	SM	
CCV44 - 1	CCV44	32342	5.3046	-7471B	PPB	31 Jan 2024 13:08:39	U	SM	
CCB44 - 1	CCB44	30	-0.0331	-7471B	PPB	31 Jan 2024 13:10:55	U	SM	