



Analytical Summary Report

Analysis Method: 1010B
Parameter: Flash Point
Run Number: LB129316
Thermometer ID: Flashpoint

Reviewed By: Nikita
Supervisor Review By: Iwona
Ambient Barometric Pressure (mmHg): 760.00
Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2745

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.4	01/31/2024	09:30
2	P1278-01		1	14	100.00	>212.0	>212.0	01/31/2024	10:00
3	P1278-01DUP		1	13	100.00	>212.0	>212.0	01/31/2024	10:30
4	P1278-03		1	12	100.00	>212.0	>212.0	01/31/2024	11:00
5	P1279-01		1	13	100.00	>212.0	>212.0	01/31/2024	11:30
6	P1282-01		1	14	100.00	>212.0	>212.0	01/31/2024	12:00
7	P1282-02		1	13	100.00	>212.0	>212.0	01/31/2024	12:30
8	P1282-03		1	12	100.00	>212.0	>212.0	01/31/2024	13:00
9	P1282-04		1	14	100.00	>212.0	>212.0	01/31/2024	13:30

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB-129316

WORKLIST(Hardcopy Internal Chain)

WorkList Name : flashpoint 01/31/24 WorkList ID : 177513 Department : Wet-Chemistry Date : 01-31-2024 09:07:57

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1278-01	AUD-1318	Water	Flash Point	Cool 4 deg C	PSEG03	D11	01/29/2024	1010B
P1278-03	AUD-1316	Water	Flash Point	Cool 4 deg C	PSEG03	D11	01/29/2024	1010B
P1279-01	1255-1259	Water	Flash Point	Cool 4 deg C	PSEG03	D11	01/29/2024	1010B
P1282-01	1240	Water	Flash Point	Cool 4 deg C	PSEG03	D11	01/29/2024	1010B
P1282-02	1241	Water	Flash Point	Cool 4 deg C	PSEG03	D11	01/29/2024	1010B
P1282-03	1242	Water	Flash Point	Cool 4 deg C	PSEG03	D11	01/29/2024	1010B
P1282-04	1243-1244	Water	Flash Point	Cool 4 deg C	PSEG03	D11	01/29/2024	1010B

Date/Time 01/31/2024 9:20
Raw Sample Received by: N.P.W.C
Raw Sample Relinquished by: N.P.W.C

Date/Time 01/31/2024 14:00
Raw Sample Received by: N.P.W.C
Raw Sample Relinquished by: N.P.W.C