



Analytical Summary Report

Analysis Method: 9040C

Parameter: Corrosivity

Run Number: LB129317

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 99.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	01/31/2024	12:25
2	CAL2	1	Water	NA	NA	20.3	7.00	01/31/2024	12:26
3	CAL3	1	Water	NA	NA	20.4	10.02	01/31/2024	12:30
4	ICV	1	Water	NA	NA	20.4	7.02	01/31/2024	12:35
5	CCV1	1	Water	NA	NA	20.3	2.01	01/31/2024	12:36
6	P1294-01	1	Water	NA	NA	20.1	7.75	01/31/2024	12:40
7	P1294-01DUP	1	Water	NA	NA	20.2	7.76	01/31/2024	12:41
8	CCV2	1	Water	NA	NA	20.3	12.02	01/31/2024	12:50

WB29317

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity p1294 WorkList ID : 177516 Department : Wet-Chemistry Date : 01-31-2024 12:06:49

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1294-01	MMW-001	Water	Corrosivity	Cool 4 deg C	ENTA05	D11	01/30/2024	9040C

SP
Date/Time 01-31-23 01:31:26 12:20
Raw Sample Received by: SP (WCI)
Raw Sample Relinquished by: SP (WCI)

Date/Time 01-31-23 13:40
Raw Sample Received by: SP (WCI)
Raw Sample Relinquished by: SP (WCI)