


LB129328

7470A

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.039	0.039	230	0.000	230					
0.2	0.200	0.180	-0.020	843	0.0 %	843	0				
2.5	2.500	2.517	0.017	10951	0.0 %	10951					
5.0	5.000	4.932	-0.068	21393	0.0 %	21393					
7.5	7.500	7.510	0.010	32545	0.0 %	32545					
10.0	10.000	10.022	0.022	43412	0.0 %	43412					

Method: 7470A

Operator: Admin

LB129328 INSTRUMENT ID : CV1

Date of Analysis: 01 Feb 2024 10:28:21

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	T
0.0 - 1	50	230	-	0.0000	7470A	PPB	01 Feb 2024 10:34:43	S	S
0.2 - 1	51.2	843	-	0.2000	7470A	PPB	01 Feb 2024 10:37:00	S	S
2.5 - 1	52.5	10951	-	2.5000	7470A	PPB	01 Feb 2024 10:39:17	S	S
5.0 - 1	53	21393	-	5.0000	7470A	PPB	01 Feb 2024 10:41:34	S	S
7.5 - 1	57.5	32545	-	7.5000	7470A	PPB	01 Feb 2024 10:43:52	S	S
10.0 - 1	5 N	43412	-	10.0000	7470A	PPB	01 Feb 2024 11:01:53	S	S
ICV88 - 1	ICV88	16276	3.7485	-	7470A	PPB	01 Feb 2024 11:04:35	U	SM
ICB88 - 1	ICB88	375	0.0722	-	7470A	PPB	01 Feb 2024 11:06:51	U	SM
CCV48 - 1	CCV48	21320	4.9146	-	7470A	PPB	01 Feb 2024 11:09:08	U	SM
CCB48 - 1	CCB48	353	0.0671	-	7470A	PPB	01 Feb 2024 11:11:23	U	SM
CRA - 1	CRA	793	0.1689	-	7470A	PPB	01 Feb 2024 11:13:40	U	SM
HighStd - 1	HighStd	42783	9.8768	-	7470A	PPB	01 Feb 2024 11:15:55	U	SM
ChkStd - 1	ChkStd	29454	6.7952	-	7470A	PPB	01 Feb 2024 11:18:10	U	SM
PB158740BL - 1	PBW	-171	-0.0540	-	7470A	PPB	01 Feb 2024 11:20:29	U	SM
PB158740BS - 1	LCSW	19098	4.4009	-	7470A	PPB	01 Feb 2024 11:22:48	U	SM
P1278-04 - 1	AUD-1316	-140	-0.0468	-	7470A	PPB	01 Feb 2024 11:25:04	U	SM
P1278-04DUP - 1	AUD-1316DUP	60	-0.0006	-	7470A	PPB	01 Feb 2024 11:27:23	U	SM
P1278-04MS - 1	AUD-1316MS	18635	4.2939	-	7470A	PPB	01 Feb 2024 11:29:39	U	SM
P1278-04MSD - 1	AUD-1316MSD	19726	4.5461	-	7470A	PPB	01 Feb 2024 11:31:57	U	SM
CCV49 - 1	CCV49	21288	4.9072	-	7470A	PPB	01 Feb 2024 11:34:15	U	SM
CCB49 - 1	CCB49	321	0.0597	-	7470A	PPB	01 Feb 2024 11:36:32	U	SM
PB158714TB - 1	PB158714TB	69	0.0015	-	7470A	PPB	01 Feb 2024 11:38:51	U	SM
P1278-04LX5 - 1		28	-0.0080	-	7470A	PPB	01 Feb 2024 11:41:06	U	SM
P1278-04A - 1		17829	4.1075	-	7470A	PPB	01 Feb 2024 11:43:22	U	SM
CCV50 - 1	CCV50	21524	4.9618	-	7470A	PPB	01 Feb 2024 11:45:38	U	SM
CCB50 - 1	CCB50	151	0.0204	-	7470A	PPB	01 Feb 2024 11:47:56	U	SM