



Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB129332

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	02/01/2024	08:25
2	CAL2	1	Water	NA	NA	20.3	6.99	02/01/2024	08:26
3	CAL3	1	Water	NA	NA	20.3	10.03	02/01/2024	08:30
4	ICV	1	Water	NA	NA	20.3	7.01	02/01/2024	08:35
5	CCV1	1	Water	NA	NA	20.3	2.01	02/01/2024	08:40
6	P1304-03	1	Water	NA	NA	22.4	7.26	02/01/2024	08:45
7	P1304-03DUP	1	Water	NA	NA	22.5	7.27	02/01/2024	08:46
8	P1304-04	1	Water	NA	NA	22.7	7.06	02/01/2024	08:50
9	P1312-01	1	Water	NA	NA	22.7	5.86	02/01/2024	09:00
10	P1312-02	1	Water	NA	NA	22.9	5.93	02/01/2024	09:05
11	CCV2	1	Water	NA	NA	20.3	12.02	02/01/2024	09:11

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph p1314 WorkList ID : 177533 Department : Wet-Chemistry Date : 02-01-2024 08:11:49

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1304-03	B BRIGHTON-MANHOLE-1-8TH-	Water	pH	Cool 4 deg C	TULL02	D11	01/30/2024	9040C
P1304-04	B BRIGHTON-MANHOLE-2-8TH-	Water	pH	Cool 4 deg C	TULL02	D11	01/30/2024	9040C
P1312-01	C NWB-1999	Water	pH	Cool 4 deg C	PSEG03	E41	01/31/2024	9040C
P1312-02	C NWW-2005	Water	pH	Cool 4 deg C	PSEG03	E41	01/31/2024	9040C

Date/Time 02/01/24 08:20
 Raw Sample Received by: JO GOC
 Raw Sample Relinquished by: CP GOC

Date/Time 02/01/24 12:40
 Raw Sample Received by: CP GOC
 Raw Sample Relinquished by: JO GOC