



Analytical Summary Report

Analysis Method: 9040C

Parameter: Corrosivity

Run Number: LB129333

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	02/01/2024	08:25
2	CAL2	1	Water	NA	NA	20.3	6.99	02/01/2024	08:26
3	CAL3	1	Water	NA	NA	20.3	10.03	02/01/2024	08:30
4	ICV	1	Water	NA	NA	20.3	7.01	02/01/2024	08:35
5	CCV1	1	Water	NA	NA	20.4	2.02	02/01/2024	09:15
6	P1314-02	1	Water	NA	NA	21.6	7.22	02/01/2024	09:25
7	P1314-02DUP	1	Water	NA	NA	21.7	7.23	02/01/2024	09:26
8	P1314-04	1	Water	NA	NA	23.4	7.23	02/01/2024	09:30
9	CCV2	1	Water	NA	NA	20.3	12.02	02/01/2024	09:33

WORKLIST(Hardcopy Internal Chain)

129333

WorkList Name : corrosivity p1314 WorkList ID : 177545 Department : Wet-Chemistry Date : 02-01-2024 08:50:23

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1314-02	TOTE-COMP	Water	Corrosivity	Cool 4 deg C	PSEG03	E41	01/31/2024	9040C
P1314-04	MOO-23-00085	Water	Corrosivity	Cool 4 deg C	PSEG03	E41	01/31/2024	9040C

Date/Time 02.01.24 09:00
 Raw Sample Received by: 30 (000)
 Raw Sample Relinquished by: 30 (000)

Date/Time 02.01.24 12:00
 Raw Sample Received by: 30 (000)
 Raw Sample Relinquished by: 30 (000)