

7470A

~~LB129341~~ LB129345


INSTRUMENT ID: CVK

nB
2/14/24

Linear

 μ Abs.:

42263



A= 0.0000e+000

B= 2.3148e-004

C= 8.7345e-003

Rho= 0.9992049

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	(%)
0.0	0.000	0.025	0.025	69	0.000	69	0				-18
0.2	0.200	0.164	-0.036	669	0.0 %	669					-2
2.5	2.500	2.440	-0.060	10505	0.0 %	10505					6
5.0	5.000	4.995	-0.005	21541	0.0 %	21541					-2
7.5	7.500	7.785	0.285	33592	0.0 %	33592					
10.0	10.000	9.792	-0.208	42263	0.0 %	42263					

Method: 7470A

Operator: Admin

Date of Analysis: 01 Feb 2024 13:46:13

LB129345

LB129341 INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	5.0	69	-	0.0000	7470A	PPB	01 Feb 2024 13:58:30	S	Std
0.2 - 1	5.2	669	-	0.2000	7470A	PPB	01 Feb 2024 14:00:48	S	Std
2.5 - 1	5.5	10505	-	2.5000	7470A	PPB	01 Feb 2024 14:03:04	S	Std
5.0 - 1	5.5	21541	-	5.0000	7470A	PPB	01 Feb 2024 14:05:21	S	Std
7.5 - 1	5.5	33592	-	7.5000	7470A	PPB	01 Feb 2024 14:07:38	S	Std
10.0 - 1	5.5	42263	-	10.0000	7470A	PPB	01 Feb 2024 14:09:58	S	Std
ICV89 - 1	ICV89	17333	4.0210	-	7470A	PPB	01 Feb 2024 14:15:28	U	SMPL
ICB89 - 1	ICB89	-125	-0.0202	-	7470A	PPB	01 Feb 2024 14:17:42	U	SMPL
CCV51 - 1	CCV51	21170	4.9091	-	7470A	PPB	01 Feb 2024 14:19:58	U	SMPL
CCB51 - 1	CCB51	-144	-0.0248	-	7470A	PPB	01 Feb 2024 14:22:14	U	SMPL
CRA - 1	CRA	954	0.2296	-	7470A	PPB	01 Feb 2024 14:24:33	U	SMPL
HighStd - 1	HighStd	41719	9.6658	-	7470A	PPB	01 Feb 2024 14:26:48	U	SMPL
ChkStd - 1	ChkStd	30538	7.0776	-	7470A	PPB	01 Feb 2024 14:29:04	U	SMPL
PB158767BL - 1	PBW	-195	-0.0364	-	7470A	PPB	01 Feb 2024 14:31:22	U	SMPL
PB158767BS - 1	LCSW	17779	4.1242	-	7470A	PPB	01 Feb 2024 14:33:40	U	SMPL
P1276-02 - 1	TP-36	325	0.0840	-	7470A	PPB	01 Feb 2024 14:35:57	U	SMPL
P1276-04 - 1	TP-37	154	0.0444	-	7470A	PPB	01 Feb 2024 14:38:16	U	SMPL
P1276-06 - 1	TP-38	6	0.0101	-	7470A	PPB	01 Feb 2024 14:40:33	U	SMPL
P1277-02 - 1	TP-39	3	0.0094	-	7470A	PPB	01 Feb 2024 14:42:51	U	SMPL
P1277-04 - 1	TP-40	15	0.0122	-	7470A	PPB	01 Feb 2024 14:45:06	U	SMPL
CCV52 - 1	CCV52	21534	4.9934	-	7470A	PPB	01 Feb 2024 14:47:21	U	SMPL
CCB52 - 1	CCB52	-203	-0.0383	-	7470A	PPB	01 Feb 2024 14:49:37	U	SMPL
P1277-06 - 1	TP-41	-16	0.0050	-	7470A	PPB	01 Feb 2024 14:51:55	U	SMPL
P1280-02 - 1	Manhole	-7	0.0071	-	7470A	PPB	01 Feb 2024 14:54:10	U	SMPL
P1280-02DUP - 1	ManholeDUP	12	0.0115	-	7470A	PPB	01 Feb 2024 14:56:26	U	SMPL
P1280-02MS - 1	ManholeMS	16807	3.8992	-	7470A	PPB	01 Feb 2024 14:58:42	U	SMPL
P1280-02MSD - 1	ManholeMSD	19148	4.4411	-	7470A	PPB	01 Feb 2024 15:00:57	U	SMPL
PB158768BL - 1	PBW	-193	-0.0359	-	7470A	PPB	01 Feb 2024 15:03:16	U	SMPL
PB158768BS - 1	LCSW	19166	4.4453	-	7470A	PPB	01 Feb 2024 15:05:35	U	SMPL
P1278-01 - 1	AUD-1318	-57	-0.0045	-	7470A	PPB	01 Feb 2024 15:07:52	U	SMPL
P1282-04 - 1	1243-1244	-43	-0.0012	-	7470A	PPB	01 Feb 2024 15:10:10	U	SMPL
P1294-01 - 1	MMW-001	28	0.0152	-	7470A	PPB	01 Feb 2024 15:12:26	U	SMPL
CCV53 - 1	CCV53	21368	4.9550	-	7470A	PPB	01 Feb 2024 15:14:41	U	SMPL
CCB53 - 1	CCB53	-188	-0.0348	-	7470A	PPB	01 Feb 2024 15:16:57	U	SMPL
P1304-03 - 1	BRIGHTON-MANHOLE-1-8TH-STREET	1172	0.2800	-	7470A	PPB	01 Feb 2024 15:19:16	U	SMPL
P1304-04 - 1	BRIGHTON-MANHOLE-2-8TH-STREET	0	0.0087	-	7470A	PPB	01 Feb 2024 15:21:31	U	SMPL
P1312-02 - 1	NWV-2005	1296	0.3087	-	7470A	PPB	01 Feb 2024 15:23:56	U	SMPL
P1312-02DUP - 1	NWV-2005DUP	1193	0.2849	-	7470A	PPB	01 Feb 2024 15:26:11	U	SMPL
P1312-02MS - 1	NWV-2005MS	16746	3.8851	-	7470A	PPB	01 Feb 2024 15:28:27	U	SMPL
P1312-02MSD - 1	NWV-2005MSD	16013	3.7154	-	7470A	PPB	01 Feb 2024 15:30:43	U	SMPL
P1280-02LX5 - 1		-118	-0.0186	-	7470A	PPB	01 Feb 2024 15:32:58	U	SMPL
P1280-02A - 1		19136	4.4383	-	7470A	PPB	01 Feb 2024 15:35:14	U	SMPL
P1312-02LX5 - 1		147	0.0428	-	7470A	PPB	01 Feb 2024 15:37:29	U	SMPL
P1312-02A - 1		19349	4.4876	-	7470A	PPB	01 Feb 2024 15:39:45	U	SMPL
CCV54 - 1	CCV54	21318	4.9434	-	7470A	PPB	01 Feb 2024 15:41:59	U	SMPL
CCB54 - 1	CCB54	-167	-0.0296	-	7470A	PPB	01 Feb 2024 15:44:15	U	SMPL
PB158699TB - 1	PB158699TB	78	0.0268	-	7470A	PPB	01 Feb 2024 15:46:30	U	SMPL
CCV55 - 1	CCV55	21265	4.9311	-	7470A	PPB	01 Feb 2024 15:48:47	U	SMPL
CCB55 - 1	CCB55	-130	-0.0214	-	7470A	PPB	01 Feb 2024 15:50:58	U	SMPL
					7470A	PPB	01 Feb 2024 16:01:02	U	SMPL
					7470A	PPB	01 Feb 2024 16:03:18	U	SMPL