

LB129350


SFAM01.1

(INSTRUMENT ID) CV1

Linear

 μ Abs.:

50709



A= 0.0000e+000

B= 1.9677e-004

C= 4.4394e-002

Rho= 0.9998482

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	o/v
0.00	0.000	0.062	0.062	91	0.000	91					1
0.05	0.050					442					1
0.20	0.200	0.176	-0.024	669	0.0 %	669	0				-12
2.50	2.500	2.523	0.023	12597	0.0 %	12597					1
5.00	5.000	4.869	-0.131	24520	0.0 %	24520					-3
7.50	7.500	7.547	0.047	38128	0.0 %	38128					1
10.0	10.000	10.022	0.022	50709	0.0 %	50709					0

LB129350

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	91	-		0 SFAM01.1	PPB	2/2/2024 9:38	Std
	0.2 S01	669	-		0.2 SFAM01.1	PPB	2/2/2024 9:40	Std
	2.5 S02	12597	-		2.5 SFAM01.1	PPB	2/2/2024 9:43	Std
	5 S03	24520	-		5 SFAM01.1	PPB	2/2/2024 9:45	Std
	7.5 S04	38128	-		7.5 SFAM01.1	PPB	2/2/2024 9:47	Std
	10 S05	50709	-		10 SFAM01.1	PPB	2/2/2024 9:49	Std
ICV090	ICV090	19758	3.9322 -		SFAM01.1	PPB	2/2/2024 9:52	SMPL
ICB090	ICB090	-200	0.005 -		SFAM01.1	PPB	2/2/2024 9:55	SMPL
CCV056	CCV056	24734	4.9113 -		SFAM01.1	PPB	2/2/2024 9:57	SMPL
CCB056	CCB056	-212	0.0027 -		SFAM01.1	PPB	2/2/2024 9:59	SMPL
PB158771BL	PBW771	3	0.045 -		SFAM01.1	PPB	2/2/2024 10:01	SMPL
P1267-01	MC0AA0	41	0.0525 -		SFAM01.1	PPB	2/2/2024 10:04	SMPL
P1267-02	MC0AA4	4	0.0452 -		SFAM01.1	PPB	2/2/2024 10:06	SMPL
P1267-03	MC0AA5	-7	0.043 -		SFAM01.1	PPB	2/2/2024 10:08	SMPL
P1267-04	MC0AA5D	-1	0.0442 -		SFAM01.1	PPB	2/2/2024 10:10	SMPL
P1267-05	MC0AA5S	5058	1.0397 -		SFAM01.1	PPB	2/2/2024 10:13	SMPL
P1267-06	MC0AA6	-1	0.0442 -		SFAM01.1	PPB	2/2/2024 10:15	SMPL
P1092-18	MC0B04	-1	0.0442 -		SFAM01.1	PPB	2/2/2024 10:17	SMPL
CCV057	CCV057	24575	4.88 -		SFAM01.1	PPB	2/2/2024 10:20	SMPL
CCB057	CCB057	-195	0.006 -		SFAM01.1	PPB	2/2/2024 10:22	SMPL