



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Nikita

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB129353

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2745

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.4	02/02/2024	08:30
2	P1284-01		1	14	100.00	>212.0	>212.0	02/02/2024	09:00
3	P1284-01DUP		1	13	100.00	>212.0	>212.0	02/02/2024	09:30
4	P1284-02		1	14	100.00	>212.0	>212.0	02/02/2024	10:00
5	P1284-03		1	12	100.00	>212.0	>212.0	02/02/2024	10:30
6	P1284-04		1	13	100.00	>212.0	>212.0	02/02/2024	11:00
7	P1284-05		1	12	100.00	>212.0	>212.0	02/02/2024	11:30
8	P1284-06		1	11	100.00	>212.0	>212.0	02/02/2024	12:00
9	P1284-07		1	13	100.00	>212.0	>212.0	02/02/2024	12:30
10	P1284-08		1	14	100.00	>212.0	>212.0	02/02/2024	13:00
11	P1284-09		1	12	100.00	>212.0	>212.0	02/02/2024	13:30

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06