



Analytical Summary Report

Analysis Method: 9045D

Parameter: Corrosivity

Run Number: LB129364

BalanceID: WC SC-4

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 99.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.4	4.01	02/02/2024	12:40
2	CAL2	1	Water	NA	NA	20.3	6.99	02/02/2024	12:41
3	CAL3	1	Water	NA	NA	20.3	10.02	02/02/2024	12:45
4	ICV	1	Water	NA	NA	20.3	7.01	02/02/2024	12:47
5	CCV1	1	Water	NA	NA	20.3	2.01	02/02/2024	12:50
6	P1066-02	1	Solid	20.02	20	22.2	10.70	02/02/2024	12:55
7	P1141-02	1	Solid	20.01	20	21.9	10.70	02/02/2024	13:00
8	P1141-04	1	Solid	20.02	20	21.8	10.49	02/02/2024	13:02
9	P1141-06	1	Solid	20.03	20	22.1	9.83	02/02/2024	13:05
10	P1141-08	1	Solid	20.05	20	22.1	8.04	02/02/2024	13:15
11	P1141-10	1	Solid	20.04	20	22.1	9.76	02/02/2024	13:20
12	P1141-12	1	Solid	20.02	20	22.3	9.92	02/02/2024	13:25
13	P1331-01	1	Solid	20.02	20	21.3	5.63	02/02/2024	13:30
14	P1337-03	1	Solid	20.03	20	23.1	6.00	02/02/2024	13:33
15	P1337-06	1	Solid	20.02	20	22.9	6.32	02/02/2024	13:35
16	CCV2	1	Water	NA	NA	20.3	12.02	02/02/2024	13:40
17	P1337-09	1	Solid	20.03	20	22.4	6.81	02/02/2024	13:45
18	P1337-12	1	Solid	20.02	20	23.1	6.82	02/02/2024	13:50
19	P1337-15	1	Solid	20.03	20	23.6	7.49	02/02/2024	14:00
20	P1337-19	1	Solid	20.04	20	23.6	6.19	02/02/2024	14:10

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
21	P1338-02	1	Solid	20.02	20	21.8	9.45	02/02/2024	14:15
22	P1338-02DUP	1	Solid	20.03	20	21.9	9.46	02/02/2024	14:16
23	CCV3	1	Water	NA	NA	20.4	2.01	02/02/2024	14:20

WORKLIST(Hardcopy Internal Chain)

129364

WorkList Name : corrosion p1337 WorkList ID : 177577 Department : Wet-Chemistry Date : 02-02-2024 10:24:11

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1066-02	WC-N-COMP	Solid	Corrosivity	Cool 4 deg C	EART12	D11	02/01/2024	9045D
P1141-02	EB-3-0-2	Solid	Corrosivity	Cool 4 deg C	HDR108	D11	01/31/2024	9045D
P1141-04	EB-4-0-2	Solid	Corrosivity	Cool 4 deg C	HDR108	D11	01/31/2024	9045D
P1141-06	EB-3-2-4	Solid	Corrosivity	Cool 4 deg C	HDR108	D11	01/31/2024	9045D
P1141-08	EB-4-6-8	Solid	Corrosivity	Cool 4 deg C	HDR108	D11	01/31/2024	9045D
P1141-10	EB-2-0-2	Solid	Corrosivity	Cool 4 deg C	HDR108	D11	01/31/2024	9045D
P1141-12	EB-2-0-2D	Solid	Corrosivity	Cool 4 deg C	HDR108	D11	01/31/2024	9045D
P1331-01	COMP-1	Solid	Corrosivity	Cool 4 deg C	HDR108	D11	01/31/2024	9045D
P1337-03	TP-1	Solid	Corrosivity	Cool 4 deg C	PSEG03	E31	02/01/2024	9045D
P1337-06	TP-2	Solid	Corrosivity	Cool 4 deg C	PSEG03	E21	02/01/2024	9045D
P1337-09	TP-3	Solid	Corrosivity	Cool 4 deg C	PSEG03	E21	02/01/2024	9045D
P1337-12	TP-4	Solid	Corrosivity	Cool 4 deg C	PSEG03	E21	02/01/2024	9045D
P1337-15	TP-5	Solid	Corrosivity	Cool 4 deg C	PSEG03	E21	02/01/2024	9045D
P1337-19	TP-6	Solid	Corrosivity	Cool 4 deg C	PSEG03	E21	02/01/2024	9045D
P1338-02	WC-S-COMP	Solid	Corrosivity	Cool 4 deg C	EART12	D11	02/01/2024	9045D

Date/Time 02/02/24 12:30
 Raw Sample Received by: SA (WC)
 Raw Sample Relinquished by: RJ (EAT-146)

Date/Time 02/02/24 17:00
 Raw Sample Received by: RJ (EAT-146)
 Raw Sample Relinquished by: SA (EAT-146)