



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB129365

Slope : 99.2

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3070
BUFFER PH 7.00 GREEN 1PINT PK6	W2981
PH 10.01 BUFFER,COLOR CD 475ML	W3043
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.4	4.01	02/02/2024	12:40
2	CAL2	1	Water	NA	NA	20.3	6.99	02/02/2024	12:41
3	CAL3	1	Water	NA	NA	20.3	10.02	02/02/2024	12:45
4	ICV	1	Water	NA	NA	20.3	7.01	02/02/2024	12:47
5	CCV1	1	Water	NA	NA	20.3	2.01	02/02/2024	12:50
6	P1066-01	1	Solid	20.02	20	22.2	10.70	02/02/2024	12:55
7	P1338-01	1	Solid	20.02	20	21.8	9.45	02/02/2024	14:15
8	P1338-01DUP	1	Solid	20.03	20	21.9	9.46	02/02/2024	14:16
9	CCV2	1	Water	NA	NA	20.3	12.02	02/02/2024	14:20

WORKLIST(Hardcopy Internal Chain)

W129365

WorkList Name : ph p1337 WorkList ID : 177591 Department : Wet-Chemistry Date : 02-02-2024 10:25:12

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1066-01	WC-N-COMP	Solid	pH	Cool 4 deg C	EART12	D11	02/01/2024	9045D
P1338-01	WC-S-COMP	Solid	pH	Cool 4 deg C	EART12	D11	02/01/2024	9045D

Date/Time 02/02/24 10:30
 Raw Sample Received by: RS CETA-WC
 Raw Sample Relinquished by: RS CETA-WC

Date/Time 02/02/24 13:30
 Raw Sample Received by: RS CETA-WC
 Raw Sample Relinquished by: RS CETA-WC