



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 2/6/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 13:15
In Date: 02/05/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:47
Out Date: 02/06/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB129367

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
P1286-01	BH6S2	1	1.18	8.56	9.74	8.2	82.0	
P1286-02	BH6S3	2	1.16	8.48	9.64	8.33	84.6	
P1286-03	BH6S4	3	1.19	8.65	9.84	7.74	75.7	
P1286-04	VHBLK001	4	1.00	1.00	2.00	2.00	100.0	vhblk
P1286-05	BH6T4	5	1.19	8.38	9.57	7.74	78.2	
P1286-06	BH6T5	6	1.13	8.69	9.82	8.43	84.0	
P1286-07	BH6T6	7	1.15	8.81	9.96	8.79	86.7	
P1286-08	BH6W9	8	1.12	8.47	9.59	7.41	74.3	
P1286-09	BH6X0	9	1.16	8.78	9.94	7.28	69.7	
P1286-10	BH6X1	10	1.15	8.40	9.55	8.18	83.7	
P1286-11	BH6X5	11	1.16	8.82	9.98	7.79	75.2	
P1286-12	BH6S5	12	1.13	8.56	9.69	7.76	77.5	
P1286-13	BH6S5MS	13	1.13	8.56	9.69	7.76	77.5	
P1286-14	BH6S5MSD	14	1.13	8.56	9.69	7.76	77.5	
P1286-15	BH6S6	15	1.19	8.70	9.89	8.52	84.3	
P1286-16	BH6S7	16	1.19	8.42	9.61	8.52	87.1	
P1286-17	BH6W0	17	1.19	8.79	9.98	7.73	74.4	
P1286-18	BH6W1	18	1.19	8.79	9.98	9.00	88.9	
P1286-19	BH6W2	19	1.18	8.62	9.8	8.29	82.5	
P1286-20	BH6S8	20	1.12	8.83	9.95	7.93	77.1	
P1286-21	BH6S9	21	1.15	8.63	9.78	8.32	83.1	
P1286-22	BH6T0	22	1.19	8.78	9.97	8.71	85.6	
P1286-23	BH6T7	23	1.13	8.73	9.86	8.25	81.6	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$