

7470A


LB129371

INSTRUMENT ID: 01

Linear

 μ Abs.:

38981



A= 0.0000e+000

B= 2.5421e-004

C= -5.7665e-003

Rho= 0.9998366

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	σ/μ
0.0	0.000	-0.010	-0.010	-17	0.000	-17	0				1
0.2	0.200	0.201	0.001	813	0.0 %	813					1
2.5	2.500	2.460	-0.040	9698	0.0 %	9698					2
5.0	5.000	5.023	0.023	19781	0.0 %	19781					2
7.5	7.500	7.623	0.123	30010	0.0 %	30010					2
10.0	10.000	9.904	-0.096	38981	0.0 %	38981					1

LB129371 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 05 Feb 2024 09:50:04

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.2	-17	-	0.0000	7470A	PPB	05 Feb 2024 09:53:35	S	Std
0.2 - 1	50.2	813	-	0.2000	7470A	PPB	05 Feb 2024 09:58:00	S	Std
2.5 - 1	55	9698	-	2.5000	7470A	PPB	05 Feb 2024 10:00:17	S	Std
5.0 - 1	55	19781	-	5.0000	7470A	PPB	05 Feb 2024 10:02:34	S	Std
7.5 - 1	55	30010	-	7.5000	7470A	PPB	05 Feb 2024 10:04:53	S	Std
10.0 - 1	55	38981	-	10.0000	7470A	PPB	05 Feb 2024 10:07:12	S	Std
ICV94 - 1	ICV94	14215	3.6079	-	7470A	PPB	05 Feb 2024 10:10:07	U	SMPL
ICB94 - 1	ICB94	-276	-0.0759	-	7470A	PPB	05 Feb 2024 10:12:33	U	SMPL
CCV67 - 1	CCV67	21146	5.3698	-	7470A	PPB	05 Feb 2024 10:14:51	U	SMPL
CCB67 - 1	CCB67	4	-0.0047	-	7470A	PPB	05 Feb 2024 10:19:44	U	SMPL
CRA - 1	CRA	831	0.2055	-	7470A	PPB	05 Feb 2024 10:21:58	U	SMPL
HighStd - 1	HighStd	39257	9.9738	-	7470A	PPB	05 Feb 2024 10:24:13	U	SMPL
ChkStd - 1	ChkStd	28094	7.1361	-	7470A	PPB	05 Feb 2024 10:26:28	U	SMPL
PB158809BL - 1	PBW	-93	-0.0294	-	7470A	PPB	05 Feb 2024 10:33:46	U	SMPL
PB158809BS - 1	LCSW	15735	3.9943	-	7470A	PPB	05 Feb 2024 10:36:03	U	SMPL
P1304-08 - 1	BRIGHTON-WEST-8TH-STREET	-206	-0.0581	-	7470A	PPB	05 Feb 2024 10:38:19	U	SMPL
P1307-02 - 1	TR-614-COMP-01	-18	-0.0103	-	7470A	PPB	05 Feb 2024 10:40:39	U	SMPL
P1307-09 - 1	TR-614-COMP-02	6	-0.0042	-	7470A	PPB	05 Feb 2024 10:42:56	U	SMPL
P1307-16 - 1	TR-614-COMP-03	21	-0.0004	-	7470A	PPB	05 Feb 2024 10:45:14	U	SMPL
P1311-02 - 1	TP-1	15	-0.0020	-	7470A	PPB	05 Feb 2024 10:47:29	U	SMPL
CCV68 - 1	CCV68	21332	5.4171	-	7470A	PPB	05 Feb 2024 10:49:45	U	SMPL
CCB68 - 1	CCB68	-266	-0.0741	-	7470A	PPB	05 Feb 2024 10:52:00	U	SMPL
P1311-04 - 1	TP-2	-26	-0.0124	-	7470A	PPB	05 Feb 2024 10:54:18	U	SMPL
P1311-06 - 1	TP-3	21	-0.0004	-	7470A	PPB	05 Feb 2024 10:56:33	U	SMPL
P1311-08 - 1	TP-4	-10	-0.0083	-	7470A	PPB	05 Feb 2024 10:58:49	U	SMPL
P1314-06 - 1	MOO-23-00071	22	-0.0002	-	7470A	PPB	05 Feb 2024 11:01:05	U	SMPL
P1314-08 - 1	MOO-23-00101	40	0.0044	-	7470A	PPB	05 Feb 2024 11:03:21	U	SMPL
P1314-10 - 1	MOO-23-00084	73	0.0128	-	7470A	PPB	05 Feb 2024 11:05:37	U	SMPL
P1314-10DUP - 1	MOO-23-00084DUP	76	0.0136	-	7470A	PPB	05 Feb 2024 11:07:54	U	SMPL
P1314-10MS - 1	MOO-23-00084MS	18508	4.6992	-	7470A	PPB	05 Feb 2024 11:10:11	U	SMPL
P1314-10MSD - 1	MOO-23-00084MSD	16345	4.1493	-	7470A	PPB	05 Feb 2024 11:15:04	U	SMPL
PB158810BL - 1	PBW	-154	-0.0449	-	7470A	PPB	05 Feb 2024 11:17:20	U	SMPL
CCV69 - 1	CCV69	19199	4.8748	-	7470A	PPB	05 Feb 2024 11:24:59	U	SMPL
CCB69 - 1	CCB69	-231	-0.0645	-	7470A	PPB	05 Feb 2024 11:27:15	U	SMPL
PB158810BS - 1	LCSW	15901	4.0365	-	7470A	PPB	05 Feb 2024 11:29:33	U	SMPL
P1314-02 - 1	TOTE-COMP	-168	-0.0485	-	7470A	PPB	05 Feb 2024 11:31:49	U	SMPL
P1314-02DUP - 1	TOTE-COMPDUP	-26	-0.0124	-	7470A	PPB	05 Feb 2024 11:34:08	U	SMPL
P1314-02MS - 1	TOTE-COMPMS	18347	4.6583	-	7470A	PPB	05 Feb 2024 11:36:23	U	SMPL
P1314-02MSD - 1	TOTE-COMPMSD	17687	4.4905	-	7470A	PPB	05 Feb 2024 11:38:39	U	SMPL
P1314-04 - 1	MOO-23-00085	-106	-0.0327	-	7470A	PPB	05 Feb 2024 11:40:57	U	SMPL
PB158743TB - 1	PB158743TB	-65	-0.0223	-	7470A	PPB	05 Feb 2024 11:43:16	U	SMPL
PB158758TB - 1	PB158758TB	-2	-0.0063	-	7470A	PPB	05 Feb 2024 11:45:33	U	SMPL
CCV70 - 1	CCV70	19612	4.9798	-	7470A	PPB	05 Feb 2024 11:47:50	U	SMPL
CCB70 - 1	CCB70	-344	-0.0932	-	7470A	PPB	05 Feb 2024 11:50:05	U	SMPL
P1314-10LX5 - 1		-80	-0.0261	-	7470A	PPB	05 Feb 2024 11:52:24	U	SMPL
P1314-10A - 1		15302	3.8842	-	7470A	PPB	05 Feb 2024 11:54:39	U	SMPL
P1314-02LX5 - 1		-217	-0.0609	-	7470A	PPB	05 Feb 2024 11:56:55	U	SMPL
P1314-02A - 1		17179	4.3613	-	7470A	PPB	05 Feb 2024 11:59:13	U	SMPL
CCV71 - 1	CCV71	20189	5.1265	-	7470A	PPB	05 Feb 2024 12:01:29	U	SMPL
CCB71 - 1	CCB71	-245	-0.0680	-	7470A	PPB	05 Feb 2024 12:03:46	U	SMPL