



Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB129375

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2745

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.4	02/05/2024	09:15
2	P1314-02		1	12	100.00	>212.0	>212.0	02/05/2024	09:45
3	P1314-02DUP		1	12	100.00	>212.0	>212.0	02/05/2024	10:15
4	P1314-04		1	14	100.00	>212.0	>212.0	02/05/2024	10:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

16129375

WorkList Name : flashpoint 02/05

WorkList ID : 177602

Department : Wet-Chemistry

Date : 02-05-2024 08:09:15

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P1314-02	TOTE-COMP	Water	Flash Point	Cool 4 deg C	PSEG03	E41	01/31/2024	1010B
P1314-04	MOO-23-00085	Water	Flash Point	Cool 4 deg C	PSEG03	E41	01/31/2024	1010B

Date/Time 02/05/2024 09:00
Raw Sample Received by: RM CMC
Raw Sample Relinquished by: ZLWCL

Date/Time 02/05/2024 11:20
Raw Sample Received by: YD CMC
Raw Sample Relinquished by: RM CMC